



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126883

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/16/2023	11:15
2	CAL2	1	Water	NA	NA	20.7	7.00	08/16/2023	11:16
3	CAL3	1	Water	NA	NA	20.7	10.02	08/16/2023	11:18
4	ICV	1	Water	NA	NA	20.7	6.98	08/16/2023	11:20
5	CCV1	1	Water	NA	NA	20.8	2.01	08/16/2023	11:22
6	O4018-01	1	Water	NA	NA	23.6	6.87	08/16/2023	11:25
7	O4018-01DUP	1	Water	NA	NA	23.7	6.88	08/16/2023	11:26
8	O4028-01	1	Water	NA	NA	22.9	4.59	08/16/2023	11:30
9	O4028-02	1	Water	NA	NA	23.0	4.68	08/16/2023	11:35
10	O4028-03	1	Water	NA	NA	22.4	4.85	08/16/2023	11:42
11	CCV2	1	Water	NA	NA	20.7	12.02	08/16/2023	11:45

WORKLIST(Hardcopy Internal Chain)

W 126483

WorkList Name : ph o4028

WorkList ID : 172757

Department : Wet-Chemistry

Date : 08-16-2023 11:04:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4018-01	ST-SUMPWATER	Water	pH	Cool 4 deg C	PSEG03	N11	08/15/2023	9040C
O4028-01	CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER13	N11	08/15/2023	9040C
O4028-02	CELL-B-POOL-WATER	Water	pH	Cool 4 deg C	STER13	N11	08/15/2023	9040C
O4028-03	CELL-C-POOL-WATER	Water	pH	Cool 4 deg C	STER13	N11	08/15/2023	9040C

11.10

Date/Time 08.16.23

Raw Sample Received by: JH (QC)

Raw Sample Relinquished by: CP SM

Date/Time

08.16.23

Raw Sample Received by: CP SM

Raw Sample Relinquished by: JH (QC)