

LB126888


7471B

INSTRUMENT ID: (V)

Linear

 μ Abs.:

66892



A= 0.0000e+000

B= 1.4830e-004

C= -4.0867e-003

Rho= 0.9997433

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	0.018	0.018	146	0.000	146					✓
0.2	0.200	0.201	0.001	1385	0.0 %	1385					✓
2.5	2.500	2.377	-0.123	16053	0.0 %	16053					✓
5.0	5.000	5.123	0.123	34573	0.0 %	34573					✓
7.5	7.500	7.565	0.065	51037	0.0 %	51037					✓
10.0	10.000	9.916	-0.084	66892	0.0 %	66892					✓

LB126888 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Aug 2023 14:41:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	146	-	0.0000	7471B	PPB	16 Aug 2023 14:43:02	S	Std
0.2 - 1	51.2	1385	-	0.2000	7471B	PPB	16 Aug 2023 14:45:18	S	Std
2.5 - 1	52.5	16053	-	2.5000	7471B	PPB	16 Aug 2023 14:47:36	S	Std
5.0 - 1	55	34573	-	5.0000	7471B	PPB	16 Aug 2023 14:49:52	S	Std
7.5 - 1	57.5	51037	-	7.5000	7471B	PPB	16 Aug 2023 14:52:09	S	Std
10.0 - 1	510	66892	-	10.0000	7471B	PPB	16 Aug 2023 14:54:24	S	Std
ICV21 - 1	ICV21	25745	3.8140	-	7471B	PPB	16 Aug 2023 14:59:40	U	SMPL
ICB21 - 1	ICB21	57	0.0044	-	7471B	PPB	16 Aug 2023 15:01:55	U	SMPL
CCV40 - 1	CCV40	34820	5.1599	-	7471B	PPB	16 Aug 2023 15:04:10	U	SMPL
CCB40 - 1	CCB40	58	0.0045	-	7471B	PPB	16 Aug 2023 15:06:25	U	SMPL
CRA - 1	CRA	1379	0.2004	-	7471B	PPB	16 Aug 2023 15:08:40	U	SMPL
HighStd - 1	HighStd	65845	9.7610	-	7471B	PPB	16 Aug 2023 15:10:55	U	SMPL
ChkStd - 1	ChkStd	46641	6.9130	-	7471B	PPB	16 Aug 2023 15:13:10	U	SMPL
PB154848BL - 1	PBS	9	-0.0028	-	7471B	PPB	16 Aug 2023 15:15:27	U	SMPL
PB154848BS - 1	LCSS	23848	3.5327	-	7471B	PPB	16 Aug 2023 15:17:43	U	SMPL
O3981-01 - 1	WC081023	24689	3.6574	-	7471B	PPB	16 Aug 2023 15:19:59	U	SMPL
O3985-01 - 1	PNNL-APM2-0811-1	5319	0.7847	-	7471B	PPB	16 Aug 2023 15:22:15	U	SMPL
O3985-03 - 1	PNNL-APM2-0811-2	8044	1.1889	-	7471B	PPB	16 Aug 2023 15:24:32	U	SMPL
O3994-01 - 1	NEVINS-SB06-Z22-23	362	0.0496	-	7471B	PPB	16 Aug 2023 15:26:49	U	SMPL
O3994-02 - 1	NEVINS-SB06-Z36-37	427	0.0592	-	7471B	PPB	16 Aug 2023 15:29:05	U	SMPL
CCV41 - 1	CCV41	34715	5.1443	-	7471B	PPB	16 Aug 2023 15:31:19	U	SMPL
CCB41 - 1	CCB41	14	-0.0020	-	7471B	PPB	16 Aug 2023 15:33:35	U	SMPL
O4003-01 - 1	TR-05-081423	1531	0.2230	-	7471B	PPB	16 Aug 2023 15:35:50	U	SMPL
O4006-01 - 1	E-1	803	0.1150	-	7471B	PPB	16 Aug 2023 15:38:05	U	SMPL
O4009-01 - 1	CONC-COMP	1414	0.2056	-	7471B	PPB	16 Aug 2023 15:40:20	U	SMPL
O4011-01 - 1	SU-03-081423	1971	0.2882	-	7471B	PPB	16 Aug 2023 15:42:36	U	SMPL
O4012-01 - 1	HD-02-081423	5131	0.7569	-	7471B	PPB	16 Aug 2023 15:44:52	U	SMPL
O4014-01 - 1	NEVINS-SB06-Z102-104	305	0.0411	-	7471B	PPB	16 Aug 2023 15:47:08	U	SMPL
O4014-01DUP - 1	NEVINS-SB06-Z102-104DUP	321	0.0435	-	7471B	PPB	16 Aug 2023 15:49:24	U	SMPL
O4014-02 - 1	O4014-01MS	24324	3.6033	-	7471B	PPB	16 Aug 2023 15:51:41	U	SMPL
O4014-03 - 1	O4014-01MSD	24748	3.6662	-	7471B	PPB	16 Aug 2023 15:53:58	U	SMPL
O4014-04 - 1	NEVINS-SB06-Z84-85	136	0.0161	-	7471B	PPB	16 Aug 2023 15:56:13	U	SMPL
CCV42 - 1	CCV42	34637	5.1327	-	7471B	PPB	16 Aug 2023 15:58:29	U	SMPL
CCB42 - 1	CCB42	50	0.0033	-	7471B	PPB	16 Aug 2023 16:00:44	U	SMPL
O4016-01 - 1	B-PP15A	1256	0.1822	-	7471B	PPB	16 Aug 2023 16:03:00	U	SMPL
O4016-02 - 1	B-PP15B	489	0.0684	-	7471B	PPB	16 Aug 2023 16:05:16	U	SMPL
O4016-03 - 1	P-PP12	3123	0.4591	-	7471B	PPB	16 Aug 2023 16:07:31	U	SMPL
O4017-01 - 1	JC-03-081423-A	7455	1.1015	-	7471B	PPB	16 Aug 2023 16:09:46	U	SMPL
O4017-03 - 1	JC-03-081423-B	5137	0.7578	-	7471B	PPB	16 Aug 2023 16:12:02	U	SMPL
O4014-01LX5 - 1		65	0.0056	-	7471B	PPB	16 Aug 2023 16:14:18	U	SMPL
O4014-01A - 1		25864	3.8317	-	7471B	PPB	16 Aug 2023 16:16:35	U	SMPL
CCV43 - 1	CCV43	34534	5.1175	-	7471B	PPB	16 Aug 2023 16:18:51	U	SMPL
CCB43 - 1	CCB43	27	-0.0001	-	7471B	PPB	16 Aug 2023 16:21:09	U	SMPL