

LB126902


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

51315



A= 0.0000e+000

B= 1.9498e-004 slope

C= -6.1597e-002 y-intercept

Rho= 0.9999332

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.00	0.000	0.025	0.025	446	0.000	446					
0.05	0.050					504	0				-16
0.20	0.200	0.169	-0.031	1181	0.0 %	1181					-1
2.50	2.500	2.469	-0.031	12977	0.0 %	12977					-1
5.00	5.000	5.026	0.026	26094	0.0 %	26094					-1
7.50	7.500	7.567	0.067	39124	0.0 %	39124					-1
10.0	10.000	9.944	-0.056	51315	0.0 %	51315					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	446	-		0 SFAM01.1	PPB	8/17/2023 13:41	Std
	0.2 S01	1181	-		0.2 SFAM01.1	PPB	8/17/2023 13:44	Std
	2.5 S02	12977	-		2.5 SFAM01.1	PPB	8/17/2023 13:48	Std
	5 S03	26094	-		5 SFAM01.1	PPB	8/17/2023 13:50	Std
	7.5 S04	39124	-		7.5 SFAM01.1	PPB	8/17/2023 13:53	Std
	10 S05	51315	-		10 SFAM01.1	PPB	8/17/2023 13:55	Std
ICV023	ICV203	22192	4.2655 -		SFAM01.1	PPB	8/17/2023 13:58	SMPL
ICB023	ICB023	385	0.0135 -		SFAM01.1	PPB	8/17/2023 14:00	SMPL
CCV048	CCV048	25849	4.9785 -		SFAM01.1	PPB	8/17/2023 14:03	SMPL
CCB048	CCB048	372	0.0109 -		SFAM01.1	PPB	8/17/2023 14:05	SMPL
PB154894BL	PBW894	62	-0.0495 -		SFAM01.1	PPB	8/17/2023 14:07	SMPL
O3865-03	CODH2	71	-0.0478 -		SFAM01.1	PPB	8/17/2023 14:12	SMPL
O3865-04	CODH4	53	-0.0513 -		SFAM01.1	PPB	8/17/2023 14:14	SMPL
O3865-05	CODH6	224	-0.0179 -		SFAM01.1	PPB	8/17/2023 14:16	SMPL
O3865-06	CODH8	327	0.0022 -		SFAM01.1	PPB	8/17/2023 14:18	SMPL
O3865-07	CODJ1	176	-0.0273 -		SFAM01.1	PPB	8/17/2023 14:21	SMPL
O3865-08	CODJ3	158	-0.0308 -		SFAM01.1	PPB	8/17/2023 14:23	SMPL
O3865-09	CODJ5	63	-0.0493 -		SFAM01.1	PPB	8/17/2023 14:25	SMPL
O3865-10	CODJ5D	47	-0.0524 -		SFAM01.1	PPB	8/17/2023 14:27	SMPL
O3865-11	CODJ5S	5110	0.9348 -		SFAM01.1	PPB	8/17/2023 14:30	SMPL
PB154895BL	PBW895	52	-0.0515 -		SFAM01.1	PPB	8/17/2023 14:32	SMPL
O3866-03	CODH3	-78	-0.0768 -		SFAM01.1	PPB	8/17/2023 14:36	SMPL
O3866-04	CODH5	-2	-0.062 -		SFAM01.1	PPB	8/17/2023 14:39	SMPL
O3866-05	CODH7	102	-0.0417 -		SFAM01.1	PPB	8/17/2023 14:41	SMPL
O3866-06	CODH9	93	-0.0435 -		SFAM01.1	PPB	8/17/2023 14:43	SMPL
O3866-07	CODJ2	26	-0.0565 -		SFAM01.1	PPB	8/17/2023 14:45	SMPL
O3866-08	CODJ4	47	-0.0524 -		SFAM01.1	PPB	8/17/2023 14:48	SMPL
O3866-09	CODJ6	9	-0.0598 -		SFAM01.1	PPB	8/17/2023 14:50	SMPL
CCV049	CCOV49	25753	4.9598 -		SFAM01.1	PPB	8/17/2023 14:52	SMPL
CCB049	CCB049	387	0.0139 -		SFAM01.1	PPB	8/17/2023 14:55	SMPL
O3866-10	CODJ6D	13	-0.0591 -		SFAM01.1	PPB	8/17/2023 14:57	SMPL
O3866-11	CODJ6S	5277	0.9673 -		SFAM01.1	PPB	8/17/2023 14:59	SMPL
O3866-12	CODJ7	35	-0.0548 -		SFAM01.1	PPB	8/17/2023 15:01	SMPL
PB154896BL	PBW896	114	-0.0394 -		SFAM01.1	PPB	8/17/2023 15:04	SMPL
O3890-01	MCOAF1	273	-0.0084 -		SFAM01.1	PPB	8/17/2023 15:06	SMPL
O3890-02	MCOAF1D	335	0.0037 -		SFAM01.1	PPB	8/17/2023 15:08	SMPL
O3890-03	MCOAF1S	5655	1.041 -		SFAM01.1	PPB	8/17/2023 15:10	SMPL
PB154897BL	PBW897	102	-0.0417 -		SFAM01.1	PPB	8/17/2023 15:13	SMPL
O3892-01	MCOAF2	169	-0.0286 -		SFAM01.1	PPB	8/17/2023 15:15	SMPL
O3892-02	MCOAF2D	117	-0.0388 -		SFAM01.1	PPB	8/17/2023 15:17	SMPL
O3892-03	MCOAF2S	5012	0.9157 -		SFAM01.1	PPB	8/17/2023 15:19	SMPL
O3865-01	CODG7	6708	1.2464 -		SFAM01.1	PPB	8/17/2023 15:24	SMPL
O3866-01	CODG8	49732	9.6353 -		SFAM01.1	PPB	8/17/2023 15:28	SMPL
CCV050	CCV050	25285	4.8686 -		SFAM01.1	PPB	8/17/2023 15:31	SMPL
CCB050	CCB050	334	0.0035 -		SFAM01.1	PPB	8/17/2023 15:33	SMPL