



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Sohil

Run Number: LB126906

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/17/2023	11:30
2	O4018-01		1	14	100.00	>212.0	>212.0	08/17/2023	12:00
3	O4018-01DUP		1	13	100.00	>212.0	>212.0	08/17/2023	12:30
4	O4030-02		1	14	100.00	>212.0	>212.0	08/17/2023	13:00
5	O4030-03		1	12	100.00	>212.0	>212.0	08/17/2023	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6626906

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-W8-15

WorkList ID : 172737

Department : Wet-Chemistry

Date : 08-15-2023 16:57:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4018-01	ST-SUMPWATER	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/15/2023	1010B
O4030-02	8410	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/16/2023	1010B
O4030-03	8395	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/16/2023	1010B

Date/Time 08/17/2023 11:20
Raw Sample Received by: RPA
Raw Sample Relinquished by: JACooley

Date/Time 08/17/2023 14:10
Raw Sample Received by: JACooley
Raw Sample Relinquished by: RPA