



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB126917

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/18/2023	11:15
2	CAL2	1	Water	NA	NA	20.7	7.00	08/18/2023	11:16
3	CAL3	1	Water	NA	NA	20.7	10.02	08/18/2023	11:19
4	ICV	1	Water	NA	NA	20.7	6.98	08/18/2023	11:20
5	CCV1	1	Water	NA	NA	20.6	2.01	08/18/2023	11:25
6	O4060-01	1	Solid	20.02	20	24.1	7.01	08/18/2023	11:30
7	O4060-01DUP	1	Solid	20.03	20	24.02	7.02	08/18/2023	11:31
8	O4075-01	1	Solid	20.05	20	20.4	9.28	08/18/2023	11:35
9	O4075-02	1	Solid	20.02	20	20.6	7.26	08/18/2023	11:40
10	CCV2	1	Water	NA	NA	20.7	12.02	08/18/2023	11:42

WORKLIST(Hardcopy Internal Chain)

WP3126917

WorkList Name : PH S O4075

WorkList ID : 172828

Department : Wet-Chemistry

Date : 08-18-2023 11:06:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4060-01	A AUD-1002	Solid	pH	Cool 4 deg C	PSEG03	N11	08/17/2023	9045D
O4075-01	P NEVINS-SB05-Z113-114	Solid	pH	Cool 4 deg C	ECOL03	N11	08/17/2023	9045D
O4075-02	P NEVINS-SB04-Z23-24	Solid	pH	Cool 4 deg C	ECOL03	N11	08/17/2023	9045D

Date/Time 08/18/23 11:10

Raw Sample Received by: 20/08/21

Raw Sample Relinquished by: CP 5m

Date/Time 08/18/23 13:00

Raw Sample Received by: CP 5m

Raw Sample Relinquished by: 20/08/21