



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126918

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/18/2023	11:15
2	CAL2	1	Water	NA	NA	20.7	7.00	08/18/2023	11:16
3	CAL3	1	Water	NA	NA	20.7	10.02	08/18/2023	11:19
4	ICV	1	Water	NA	NA	20.7	6.98	08/18/2023	11:20
5	CCV1	1	Water	NA	NA	20.6	2.02	08/18/2023	11:45
6	O4030-02	1	Water	NA	NA	27.0	8.03	08/18/2023	11:50
7	O4030-03	1	Water	NA	NA	26.3	5.99	08/18/2023	11:52
8	O4060-02	1	Water	NA	NA	26.6	6.75	08/18/2023	12:00
9	O4072-01	1	Water	NA	NA	26.5	7.10	08/18/2023	12:10
10	O4074-01	1	Water	NA	NA	20.6	7.65	08/18/2023	12:15
11	O4074-01DUP	1	Water	NA	NA	20.6	7.66	08/18/2023	12:16
12	CCV2	1	Water	NA	NA	20.8	12.03	08/18/2023	12:20

WORKLIST(Hardcopy Internal Chain)

126918

WorkList Name : PH 04067

WorkList ID : 172827

Department : Wet-Chemistry

Date : 08-18-2023 11:05:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4030-02	8410	Water	pH	Cool 4 deg C	PSEG03	N11	08/16/2023	9040C
O4030-03	8395	Water	pH	Cool 4 deg C	PSEG03	N11	08/16/2023	9040C
O4060-02	AUD-1038	Water	pH	Cool 4 deg C	PSEG03	N11	08/17/2023	9040C
O4072-01	50112	Water	pH	Cool 4 deg C	PSEG03	N11	08/17/2023	9040C
O4074-01	MAPLE-SHADE-COMP	Water	pH	Cool 4 deg C	PSEG03	N11	08/17/2023	9040C

Date/Time 08-18-23 11:10

Raw Sample Received by: 7016001

Raw Sample Relinquished by: CP SM

Date/Time 08-18-23 13:30

Raw Sample Received by: CP SM

Raw Sample Relinquished by: 7016001