



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB126929

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	08/18/2023	09:10
2	O4060-02		1	12	100.00	>212.0	>212.0	08/18/2023	09:40
3	O4060-02DUP		1	13	100.00	>212.0	>212.0	08/18/2023	10:10
4	O4072-01		1	14	100.00	>212.0	>212.0	08/18/2023	10:40
5	O4074-01		1	14	100.00	>212.0	>212.0	08/18/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6126929

WorkList Name : FP 8-17

WorkList ID : 172808

Department : Wet-Chemistry

Date : 08-17-2023 17:41:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4060-02	AUD-1038	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/17/2023	1010B
O4072-01	50112	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/17/2023	1010B
O4074-01	MAPLE-SHADE-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/17/2023	1010B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/18/2023 09:00
PR1
286204

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/18/2023 11:45
786204
PR1