



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/21/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 08/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 08/20/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126939

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4084-01	GCEF2	1	1.1	8.78	9.88	8.86	88.4	
O4084-02	GCEF3	2	1.19	8.69	9.88	8.57	84.9	
O4084-03	GCEF3D	3	1.19	8.69	9.88	8.57	84.9	
O4084-04	GCEF3S	4	1.19	8.69	9.88	8.57	84.9	
O4084-05	GCEF4	5	1.19	8.79	9.98	9.47	94.2	
O4084-06	GCEF5	6	1.15	8.82	9.97	8.83	87.1	
O4084-07	GCEF6	7	1.15	8.69	9.84	8.02	79.1	
O4084-08	GCEF7	8	1.13	8.38	9.51	8.04	82.5	
O4084-09	GCEF8	9	1.16	8.41	9.57	8.37	85.7	
O4084-10	GCEF9	10	1.16	8.39	9.55	7.67	77.6	
O4084-11	GCEG0	11	1.18	8.47	9.65	8.49	86.3	
O4084-12	GCEG1	12	1.15	8.82	9.97	8.68	85.4	
O4084-13	GCEG2	13	1.19	8.73	9.92	8.27	81.1	
O4084-14	GCEG3	14	1.15	8.82	9.97	8.44	82.7	
O4084-15	GCEG4	15	1.19	8.51	9.7	7.39	72.9	
O4084-16	GCEG5	16	1.11	8.66	9.77	7.75	76.7	
O4084-17	GCEG6	17	1.11	8.55	9.66	8.78	89.7	
O4084-18	GCEG7	18	1.19	8.66	9.85	8.51	84.5	
O4084-19	GCEG8	19	1.16	8.82	9.98	8.21	79.9	
O4084-20	GCEG9	20	1.19	8.45	9.64	8.48	86.3	
O4084-21	GCEH0	21	1.14	8.80	9.94	8.87	87.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$