

LB126942


7471B

INSTRUMENT ID: CVJ

Linear

μ Abs.:

54751



A= 0.0000e+000

B= 1.7552e-004

C= -6.6855e-002

Rho= 0.9968057

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.010	-0.010	323	0.000	323					0/11
0.2	0.200	0.204	0.004	1543	0.0 %	1543					2
2.5	2.500	2.297	-0.203	13470	0.0 %	13470					-8
5.0	5.000	5.176	0.176	29868	0.0 %	29868					4
7.5	7.500	7.990	0.490	45904	0.0 %	45904					7
10.0	10.000	9.543	-0.457	54751	0.0 %	54751					-5

LB126942 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Aug 2023 10:17:14

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	323	-	0.0000	7471B	PPB	21 Aug 2023 11:49:00	S	Std
0.2 - 1	50.2	1543	-	0.2000	7471B	PPB	21 Aug 2023 11:51:17	S	Std
2.5 - 1	52.5	13470	-	2.5000	7471B	PPB	21 Aug 2023 11:53:33	S	Std
5.0 - 1	55	29868	-	5.0000	7471B	PPB	21 Aug 2023 11:58:26	S	Std
7.5 - 1	57.5	45904	-	7.5000	7471B	PPB	21 Aug 2023 12:03:16	S	Std
10.0 - 1	510	54751	-	10.0000	7471B	PPB	21 Aug 2023 12:08:20	S	Std
ICV25 - 1	ICV25	23101	3.9878	-	7471B	PPB	21 Aug 2023 12:13:49	U	SMPL
ICB25 - 1	ICB25	269	-0.0196	-	7471B	PPB	21 Aug 2023 12:16:04	U	SMPL
CCV55 - 1	CCV55	29582	5.1254	-	7471B	PPB	21 Aug 2023 12:20:55	U	SMPL
CCB55 - 1	CCB55	245	-0.0239	-	7471B	PPB	21 Aug 2023 12:23:10	U	SMPL
CRA - 1	CRA	1550	0.2052	-	7471B	PPB	21 Aug 2023 12:25:24	U	SMPL
HighStd - 1	HighStd	55103	9.6048	-	7471B	PPB	21 Aug 2023 12:27:39	U	SMPL
ChkStd - 1	ChkStd	39169	6.8081	-	7471B	PPB	21 Aug 2023 12:32:41	U	SMPL
PB154940BL - 1	PBS	238	-0.0251	-	7471B	PPB	21 Aug 2023 12:34:57	U	SMPL
PB154940BS - 1	LCSS	22033	3.8004	-	7471B	PPB	21 Aug 2023 12:37:13	U	SMPL
O4056-04 - 1	P001-SS004-0002-01	9232	1.5535	-	7471B	PPB	21 Aug 2023 12:39:29	U	SMPL
O4056-05 - 1	P001-SS009-0002-01	4736	0.7644	-	7471B	PPB	21 Aug 2023 12:41:46	U	SMPL
O4056-06 - 1	P001-SS010-0002-01	4173	0.6656	-	7471B	PPB	21 Aug 2023 12:44:03	U	SMPL
O4056-07 - 1	P001-SS011-0002-01	4969	0.8053	-	7471B	PPB	21 Aug 2023 12:46:19	U	SMPL
O4056-08 - 1	P001-SS012-0002-01	4835	0.7818	-	7471B	PPB	21 Aug 2023 12:48:34	U	SMPL
CCV56 - 1	CCV56	29555	5.1206	-	7471B	PPB	21 Aug 2023 12:50:49	U	SMPL
CCB56 - 1	CCB56	248	-0.0233	-	7471B	PPB	21 Aug 2023 12:53:04	U	SMPL
O4056-09 - 1	P001-SS013-0002-01	3234	0.5008	-	7471B	PPB	21 Aug 2023 12:55:19	U	SMPL
O4056-10 - 1	P001-SS014-0002-01	5614	0.9185	-	7471B	PPB	21 Aug 2023 12:57:34	U	SMPL
O4056-11 - 1	P001-SS015-0002-01	5025	0.8151	-	7471B	PPB	21 Aug 2023 12:59:49	U	SMPL
O4056-12 - 1	P001-SS016-0002-01	8653	1.4519	-	7471B	PPB	21 Aug 2023 13:02:04	U	SMPL
O4056-13 - 1	P001-SS017-0002-01	8187	1.3701	-	7471B	PPB	21 Aug 2023 13:04:20	U	SMPL
O4056-14 - 1	P001-SS019-0001-01	11912	2.0239	-	7471B	PPB	21 Aug 2023 13:06:36	U	SMPL
O4056-14DUP - 1	P001-SS019-0001-01DUP	12083	2.0539	-	7471B	PPB	21 Aug 2023 13:08:52	U	SMPL
O4056-15 - 1	O4056-14MS	37076	6.4407	-	7471B	PPB	21 Aug 2023 13:11:09	U	SMPL
O4056-16 - 1	O4056-14MSD	35800	6.2168	-	7471B	PPB	21 Aug 2023 13:13:26	U	SMPL
O4056-17 - 1	P001-SS019-0001-02	11515	1.9543	-	7471B	PPB	21 Aug 2023 13:15:41	U	SMPL
CCV57 - 1	CCV57	29876	5.1770	-	7471B	PPB	21 Aug 2023 13:17:57	U	SMPL
CCB57 - 1	CCB57	229	-0.0267	-	7471B	PPB	21 Aug 2023 13:20:12	U	SMPL
O4056-18 - 1	P001-SS022-0002-01	8990	1.5111	-	7471B	PPB	21 Aug 2023 13:22:28	U	SMPL
O4056-19 - 1	P001-SS023-0002-01	37971	6.5978	-	7471B	PPB	21 Aug 2023 13:24:43	U	SMPL
O4056-20 - 1	P001-SS024-0002-01	145541	25.4785	-	7471B	PPB	21 Aug 2023 13:26:58	U	SMPL
O4056-21 - 1	P001-SS025-0002-01	4910	0.7949	-	7471B	PPB	21 Aug 2023 13:29:13	U	SMPL
O4056-22 - 1	P001-SS026-0002-01	12034	2.0453	-	7471B	PPB	21 Aug 2023 13:31:32	U	SMPL
O4056-23 - 1	P001-SS027-0002-01	8540	1.4321	-	7471B	PPB	21 Aug 2023 13:33:48	U	SMPL
PB154941BL - 1	PBS	305	-0.0133	-	7471B	PPB	21 Aug 2023 13:36:05	U	SMPL
PB154941BS - 1	LCSS	25339	4.3806	-	7471B	PPB	21 Aug 2023 13:38:21	U	SMPL
O4057-01 - 1	P001-SS028-0002-01	1868	0.2610	-	7471B	PPB	21 Aug 2023 13:40:38	U	SMPL
O4057-02 - 1	P001-SS031-0002-01	4468	0.7174	-	7471B	PPB	21 Aug 2023 13:42:54	U	SMPL
CCV58 - 1	CCV58	29347	5.0841	-	7471B	PPB	21 Aug 2023 13:45:09	U	SMPL
CCB58 - 1	CCB58	248	-0.0233	-	7471B	PPB	21 Aug 2023 13:47:24	U	SMPL
O4057-03 - 1	P001-SS033-0002-01	6782	1.1235	-	7471B	PPB	21 Aug 2023 13:49:40	U	SMPL
O4057-04 - 1	P001-SS034-0002-01	4825	0.7800	-	7471B	PPB	21 Aug 2023 13:51:56	U	SMPL
O4057-05 - 1	P001-SS035-0002-01	3195	0.4939	-	7471B	PPB	21 Aug 2023 13:54:11	U	SMPL
O4057-06 - 1	P001-SS040-0002-01	12370	2.1043	-	7471B	PPB	21 Aug 2023 13:56:26	U	SMPL
O4057-07 - 1	P001-SS042-0002-01	3434	0.5359	-	7471B	PPB	21 Aug 2023 13:58:42	U	SMPL
O4057-08 - 1	P001-SS047-0002-01	9450	1.5918	-	7471B	PPB	21 Aug 2023 14:00:58	U	SMPL
O4057-09 - 1	P001-SS049-0002-01	13204	2.2507	-	7471B	PPB	21 Aug 2023 14:03:15	U	SMPL
O4057-10 - 1	P001-SS053-0002-01	11495	1.9507	-	7471B	PPB	21 Aug 2023 14:05:32	U	SMPL
O4057-11 - 1	P001-SS056-0002-01	3007	0.4609	-	7471B	PPB	21 Aug 2023 14:07:48	U	SMPL
O4057-12 - 1	P001-SS058-0002-01	31414	5.4469	-	7471B	PPB	21 Aug 2023 14:10:04	U	SMPL
CCV59 - 1	CCV59	29659	5.1389	-	7471B	PPB	21 Aug 2023 14:12:20	U	SMPL
CCB59 - 1	CCB59	262	-0.0209	-	7471B	PPB	21 Aug 2023 14:14:36	U	SMPL
O4057-13 - 1	P001-SS061-0002-01	8371	1.4024	-	7471B	PPB	21 Aug 2023 14:16:51	U	SMPL
O4057-14 - 1	P001-SS062-0002-01	10516	1.7789	-	7471B	PPB	21 Aug 2023 14:19:06	U	SMPL
O4057-15 - 1	P001-SS064-0002-01	4922	0.7971	-	7471B	PPB	21 Aug 2023 14:21:21	U	SMPL
O4057-16 - 1	P001-SS065-0002-01	18571	3.1927	-	7471B	PPB	21 Aug 2023 14:23:37	U	SMPL
O4057-17 - 1	P001-SS066-0002-01	10423	1.7626	-	7471B	PPB	21 Aug 2023 14:25:52	U	SMPL
O4057-17DUP - 1	P001-SS066-0002-01DUP	9048	1.5212	-	7471B	PPB	21 Aug 2023 14:28:08	U	SMPL
O4057-18 - 1	O4057-17MS	33958	5.8934	-	7471B	PPB	21 Aug 2023 14:33:10	U	SMPL
O4057-19 - 1	O4057-17MSD	32293	5.6012	-	7471B	PPB	21 Aug 2023 14:35:27	U	SMPL
O4057-20 - 1	P001-SS066-0002-02	10682	1.8080	-	7471B	PPB	21 Aug 2023 14:37:43	U	SMPL
O4057-21 - 1	P001-SS067-0002-01	7262	1.2078	-	7471B	PPB	21 Aug 2023 14:39:59	U	SMPL
CCV60 - 1	CCV60	29526	5.1155	-	7471B	PPB	21 Aug 2023 14:42:15	U	SMPL
CCB60 - 1	CCB60	236	-0.0254	-	7471B	PPB	21 Aug 2023 14:44:31	U	SMPL
O4057-22 - 1	P001-SS068-0002-01	2208	0.3207	-	7471B	PPB	21 Aug 2023 14:46:48	U	SMPL
O4057-23 - 1	P001-SS071-0002-01	4133	0.6586	-	7471B	PPB	21 Aug 2023 14:49:03	U	SMPL
PB154942BL - 1	PBS	186	-0.0342	-	7471B	PPB	21 Aug 2023 14:51:19	U	SMPL
PB154942BS - 1	LCSS	24732	4.2741	-	7471B	PPB	21 Aug 2023 14:58:55	U	SMPL
O4046-01 - 1	SB-02(0-5)	6434	1.0624	-	7471B	PPB	21 Aug 2023 15:01:10	U	SMPL
O4046-01DUP - 1	SB-02(0-5)DUP	6893	1.1430	-	7471B	PPB	21 Aug 2023 15:03:26	U	SMPL
O4046-01MS - 1	SB-02(0-5)MS	30699	5.3214	-	7471B	PPB	21 Aug 2023 15:05:43	U	SMPL
O4046-01MSD - 1	SB-02(0-5)MSD	30365	5.2628	-	7471B	PPB	21 Aug 2023 15:07:59	U	SMPL
O4046-02 - 1	SB-06(0-5)	14571	2.4906	-	7471B	PPB	21 Aug 2023 15:10:16	U	SMPL
O4046-03 - 1	SB-01(0-5)	4259	0.6807	-	7471B	PPB	21 Aug 2023 15:12:32	U	SMPL
CCV61 - 1	CCV61	29247	5.0666	-	7471B	PPB	21 Aug 2023 15:26:15	U	SMPL

LB126942 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Aug 2023 10:17:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB61 - 1	CCB61	262	-0.0209		-7471B	PPB	21 Aug 2023 15:28:31	U	SMPL
O4046-04 - 1	SB-07(0-5)	17529	3.0098		-7471B	PPB	21 Aug 2023 15:30:46	U	SMPL
O4047-01 - 1	NEVINS-SB05-Z26-27	479	0.0172		-7471B	PPB	21 Aug 2023 15:33:03	U	SMPL
O4047-02 - 1	NEVINS-SB05-Z60-61	446	0.0114		-7471B	PPB	21 Aug 2023 15:35:19	U	SMPL
O4047-03 - 1	NEVINS-SB05-Z60-61-Q	512	0.0230		-7471B	PPB	21 Aug 2023 15:37:35	U	SMPL
O4047-04 - 1	NEVINS-SB05-Z83-84	465	0.0148		-7471B	PPB	21 Aug 2023 15:39:51	U	SMPL
O4058-01 - 1	SU-04-081723	2906	0.4432		-7471B	PPB	21 Aug 2023 15:42:08	U	SMPL
O4060-01 - 1	AUD-1002	852	0.0827		-7471B	PPB	21 Aug 2023 15:44:24	U	SMPL
O4070-01 - 1	WC-1	4269	0.6824		-7471B	PPB	21 Aug 2023 15:46:40	U	SMPL
O4075-01 - 1	NEVINS-SB05-Z113-114	315	-0.0116		-7471B	PPB	21 Aug 2023 15:48:57	U	SMPL
O4075-02 - 1	NEVINS-SB04-Z23-24	703	0.0565		-7471B	PPB	21 Aug 2023 15:51:14	U	SMPL
CCV62 - 1	CCV62	30154	5.2258		-7471B	PPB	21 Aug 2023 15:55:57	U	SMPL
CCB62 - 1	CCB62	226	-0.0272		-7471B	PPB	21 Aug 2023 15:58:14	U	SMPL
O4056-14LX5 - 1		2655	0.3991		-7471B	PPB	21 Aug 2023 16:03:16	U	SMPL
O4056-14A - 1		35896	6.2336		-7471B	PPB	21 Aug 2023 16:05:32	U	SMPL
O4057-17LX5 - 1		1947	0.2749		-7471B	PPB	21 Aug 2023 16:07:48	U	SMPL
O4057-17A - 1		35966	6.2459		-7471B	PPB	21 Aug 2023 16:10:04	U	SMPL
O4046-01LX5 - 1		1412	0.1810		-7471B	PPB	21 Aug 2023 16:15:12	U	SMPL
O4046-01A - 1		30809	5.3407		-7471B	PPB	21 Aug 2023 16:17:28	U	SMPL
O4056-20DLX5 - 1		23340	4.0298		-7471B	PPB	21 Aug 2023 16:19:45	U	SMPL
CCV63 - 1	CCV63	28235	4.8889		-7471B	PPB	21 Aug 2023 16:24:42	U	SMPL
CCB63 - 1	CCB63	214	-0.0293		-7471B	PPB	21 Aug 2023 16:26:59	U	SMPL