

LB126957


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

64293



A= 0.0000e+000

B= 1.5652e-004

C= -4.9656e-003

Rho= 0.9998896

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.046	0.046	323	0.000	323					-
0.2	0.200	0.194	-0.006	1269	0.0 %	1269					-3
2.5	2.500	2.453	-0.047	15701	0.0 %	15701					-2
5.0	5.000	5.042	0.042	32245	0.0 %	32245					-1
7.5	7.500	7.408	-0.092	47359	0.0 %	47359					-1
10.0	10.000	10.058	0.058	64293	0.0 %	64293					-1

LB126951 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Aug 2023 08:28:51

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	323	-	0.0000	7471B	PPB	22 Aug 2023 09:22:17	S	Std
0.2 - 1	502	1269	-	0.2000	7471B	PPB	22 Aug 2023 09:24:34	S	Std
2.5 - 1	52.5	15701	-	2.5000	7471B	PPB	22 Aug 2023 09:26:50	S	Std
5.0 - 1	55	32245	-	5.0000	7471B	PPB	22 Aug 2023 09:29:07	S	Std
7.5 - 1	57.5	47359	-	7.5000	7471B	PPB	22 Aug 2023 09:31:23	S	Std
10.0 - 1	51v	64293	-	10.0000	7471B	PPB	22 Aug 2023 09:37:44	S	Std
ICV26 - 1	ICV26	23286	3.6398	-	7471B	PPB	22 Aug 2023 09:40:35	U	SMPL
ICB26 - 1	ICB26	-46	-0.0122	-	7471B	PPB	22 Aug 2023 09:42:50	U	SMPL
CCV64 - 1	CCV64	32200	5.0350	-	7471B	PPB	22 Aug 2023 09:45:06	U	SMPL
CCB64 - 1	CCB64	73	0.0065	-	7471B	PPB	22 Aug 2023 09:47:21	U	SMPL
CRA - 1	CRA	1275	0.1946	-	7471B	PPB	22 Aug 2023 09:49:36	U	SMPL
HighStd - 1	HighStd	65218	10.2031	-	7471B	PPB	22 Aug 2023 09:51:50	U	SMPL
ChkStd - 1	ChkStd	43198	6.7565	-	7471B	PPB	22 Aug 2023 09:54:05	U	SMPL
PB154991BL - 1	PBS	364	0.0520	-	7471B	PPB	22 Aug 2023 09:56:23	U	SMPL
PB154991BS - 1	LCSS	24602	3.8458	-	7471B	PPB	22 Aug 2023 09:58:39	U	SMPL
O4086-01 - 1	B-PP06A	2858	0.4424	-	7471B	PPB	22 Aug 2023 10:00:55	U	SMPL
O4086-01DUP - 1	B-PP06ADUP	3128	0.4846	-	7471B	PPB	22 Aug 2023 10:03:11	U	SMPL
O4086-01MS - 1	B-PP06AMS	28442	4.4468	-	7471B	PPB	22 Aug 2023 10:08:47	U	SMPL
O4086-01MSD - 1	B-PP06AMSD	24742	3.8677	-	7471B	PPB	22 Aug 2023 10:11:03	U	SMPL
O4086-02 - 1	B-PP06B	12757	1.9918	-	7471B	PPB	22 Aug 2023 10:13:18	U	SMPL
CCV65 - 1	CCV65	31652	4.9493	-	7471B	PPB	22 Aug 2023 10:15:35	U	SMPL
CCB65 - 1	CCB65	84	0.0082	-	7471B	PPB	22 Aug 2023 10:17:50	U	SMPL
O4086-03 - 1	B-PP07A	786	0.1181	-	7471B	PPB	22 Aug 2023 10:20:06	U	SMPL
O4086-04 - 1	B-PP07B	985	0.1492	-	7471B	PPB	22 Aug 2023 10:22:21	U	SMPL
O4086-05 - 1	B-PP09A	2083	0.3211	-	7471B	PPB	22 Aug 2023 10:24:36	U	SMPL
O4086-06 - 1	B-PP09B	1615	0.2478	-	7471B	PPB	22 Aug 2023 10:26:52	U	SMPL
O4086-07 - 1	B-PP09C	362	0.0517	-	7471B	PPB	22 Aug 2023 10:29:07	U	SMPL
O4086-08 - 1	B-PP13A	6194	0.9645	-	7471B	PPB	22 Aug 2023 10:31:23	U	SMPL
O4086-09 - 1	B-PP13B	1094	0.1663	-	7471B	PPB	22 Aug 2023 10:33:40	U	SMPL
O4086-11 - 1	B-PP14A	2773	0.4291	-	7471B	PPB	22 Aug 2023 10:35:57	U	SMPL
O4086-12 - 1	B-PP14B	573	0.0847	-	7471B	PPB	22 Aug 2023 10:38:13	U	SMPL
O4086-13 - 1	B-PP17A	1591	0.2441	-	7471B	PPB	22 Aug 2023 10:40:29	U	SMPL
CCV66 - 1	CCV66	31886	4.9859	-	7471B	PPB	22 Aug 2023 10:42:44	U	SMPL
CCB66 - 1	CCB66	125	0.0146	-	7471B	PPB	22 Aug 2023 10:45:00	U	SMPL
O4086-14 - 1	B-PP17B	1109	0.1686	-	7471B	PPB	22 Aug 2023 10:47:15	U	SMPL
O4086-15 - 1	B-PP18A	1677	0.2575	-	7471B	PPB	22 Aug 2023 10:49:31	U	SMPL
O4086-16 - 1	B-PP18B	501	0.0735	-	7471B	PPB	22 Aug 2023 10:51:46	U	SMPL
O4086-18 - 1	DUP06	497	0.0728	-	7471B	PPB	22 Aug 2023 10:54:04	U	SMPL
O4086-19 - 1	B-PP10A	3572	0.5541	-	7471B	PPB	22 Aug 2023 10:56:21	U	SMPL
O4086-20 - 1	B-PP10B	1257	0.1918	-	7471B	PPB	22 Aug 2023 10:58:38	U	SMPL
O4087-01 - 1	EO-03-081823	3512	0.5447	-	7471B	PPB	22 Aug 2023 11:00:55	U	SMPL
O4089-01 - 1	BU-3-081823	1595	0.2447	-	7471B	PPB	22 Aug 2023 11:03:12	U	SMPL
PB154992BL - 1	PBS	150	0.0185	-	7471B	PPB	22 Aug 2023 11:05:29	U	SMPL
PB154992BS - 1	LCSS	24968	3.9031	-	7471B	PPB	22 Aug 2023 11:07:45	U	SMPL
CCV67 - 1	CCV67	32572	5.0933	-	7471B	PPB	22 Aug 2023 11:10:00	U	SMPL
CCB67 - 1	CCB67	192	0.0251	-	7471B	PPB	22 Aug 2023 11:12:16	U	SMPL
O4080-01 - 1	ETGI-340	2239	0.3455	-	7471B	PPB	22 Aug 2023 11:14:31	U	SMPL
O4082-01 - 1	277	37798	5.9113	-	7471B	PPB	22 Aug 2023 11:16:47	U	SMPL
O4082-03 - 1	364	32956	5.1534	-	7471B	PPB	22 Aug 2023 11:19:02	U	SMPL
O4082-05 - 1	72-11929	5999	0.9340	-	7471B	PPB	22 Aug 2023 11:21:18	U	SMPL
O4082-07 - 1	217	6056	0.9429	-	7471B	PPB	22 Aug 2023 11:23:34	U	SMPL
O4082-09 - 1	362	9441	1.4728	-	7471B	PPB	22 Aug 2023 11:25:50	U	SMPL
O4090-01 - 1	TP-I	615	0.0913	-	7471B	PPB	22 Aug 2023 11:28:07	U	SMPL
O4090-04 - 1	TP-K	2254	0.3478	-	7471B	PPB	22 Aug 2023 11:30:24	U	SMPL
O4090-07 - 1	TP-L	1444	0.2211	-	7471B	PPB	22 Aug 2023 11:32:41	U	SMPL
O4090-10 - 1	TP-N	8963	1.3979	-	7471B	PPB	22 Aug 2023 11:34:57	U	SMPL
CCV68 - 1	CCV68	32880	5.1415	-	7471B	PPB	22 Aug 2023 11:37:12	U	SMPL
CCB68 - 1	CCB68	212	0.0282	-	7471B	PPB	22 Aug 2023 11:39:27	U	SMPL
O4091-01 - 1	TP-A	2975	0.4607	-	7471B	PPB	22 Aug 2023 11:41:42	U	SMPL
O4091-04 - 1	TP-B	990	0.1500	-	7471B	PPB	22 Aug 2023 11:43:58	U	SMPL
O4091-07 - 1	TP-C	1165	0.1774	-	7471B	PPB	22 Aug 2023 11:46:13	U	SMPL
O4091-10 - 1	TP-E	2213	0.3414	-	7471B	PPB	22 Aug 2023 11:48:28	U	SMPL
O4091-13 - 1	TP-H	772	0.1159	-	7471B	PPB	22 Aug 2023 11:50:44	U	SMPL
O4091-16 - 1	TP-G	901	0.1361	-	7471B	PPB	22 Aug 2023 11:53:01	U	SMPL
O4091-19 - 1	TP-F	1989	0.3064	-	7471B	PPB	22 Aug 2023 11:55:18	U	SMPL
O4091-19DUP - 1	TP-FDUP	1792	0.2755	-	7471B	PPB	22 Aug 2023 11:57:35	U	SMPL
O4091-19MS - 1	TP-FMS	24279	3.7952	-	7471B	PPB	22 Aug 2023 11:59:52	U	SMPL
O4091-19MSD - 1	TP-FMSD	25067	3.9186	-	7471B	PPB	22 Aug 2023 12:02:07	U	SMPL
CCV69 - 1	CCV69	32841	5.1354	-	7471B	PPB	22 Aug 2023 12:04:23	U	SMPL
CCB69 - 1	CCB69	203	0.0268	-	7471B	PPB	22 Aug 2023 12:06:39	U	SMPL
O4100-01 - 1	WC-1	14470	2.2599	-	7471B	PPB	22 Aug 2023 12:08:55	U	SMPL
O4086-01LX5 - 1		504	0.0739	-	7471B	PPB	22 Aug 2023 12:11:11	U	SMPL
O4086-01A - 1		28525	4.4598	-	7471B	PPB	22 Aug 2023 12:13:26	U	SMPL

LB126951 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Aug 2023 08:28:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
O4091-19LX5 - 1		333	0.0472		-7471B	PPB	22 Aug 2023 12:15:42	U	SMPL
O4091-19A - 1		22490	3.5152		-7471B	PPB	22 Aug 2023 12:17:58	U	SMPL
CCV70 - 1	CCV70	32827	5.1332		-7471B	PPB	22 Aug 2023 12:20:15	U	SMPL
CCB70 - 1	CCB70	243	0.0331		-7471B	PPB	22 Aug 2023 12:22:31	U	SMPL