



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB126952

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/22/2023	08:05
2	CAL2	1	Water	NA	NA	20.7	7.00	08/22/2023	08:06
3	CAL3	1	Water	NA	NA	20.7	10.02	08/22/2023	08:10
4	ICV	1	Water	NA	NA	20.6	6.99	08/22/2023	08:14
5	CCV1	1	Water	NA	NA	20.6	2.01	08/22/2023	08:15
6	O4089-01	1	Solid	20.02	20	21.0	8.20	08/22/2023	08:20
7	O4089-01DUP	1	Solid	20.03	20	21.1	8.21	08/22/2023	08:21
8	CCV2	1	Water	NA	NA	20.6	12.02	08/21/2023	08:25

UB 126952

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O4089

WorkList ID : 172905

Department : Wet-Chemistry

Date : 08-22-2023 07:42:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4089-01	BU-3-081823	Solid	pH	Cool 4 deg C	PSEG05	N11	08/18/2023	9045D

Date/Time 08/22/23

Raw Sample Received by: JOC

Raw Sample Relinquished by: JOC

Date/Time 08/22/23

Raw Sample Received by: JOC

Raw Sample Relinquished by: JOC