

LB126959


7470A

INSTRUMENT 115. CV1

Linear

 μ Abs.:

56423



A= 0.0000e+000

B= 1.7558e-004

C= 5.4845e-002

Rho= 0.9995467

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/100
0.0	0.000	0.065	0.065	59	0.000	59	0				1
0.2	0.200	0.228	0.028	989	0.0 %	989					14
2.5	2.500	2.447	-0.053	13623	0.0 %	13623					-2
5.0	5.000	4.819	-0.181	27133	0.0 %	27133					-4
7.5	7.500	7.679	0.179	43420	0.0 %	43420					2
10.0	10.000	9.962	-0.038	56423	0.0 %	56423					0

LB126959 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Aug 2023 12:37:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	59	-	0.0000	7470A	PPB	22 Aug 2023 12:39:10	S	Std
0.2 - 1	50.2	989	-	0.2000	7470A	PPB	22 Aug 2023 12:41:27	S	Std
2.5 - 1	52.5	13623	-	2.5000	7470A	PPB	22 Aug 2023 12:43:44	S	Std
5.0 - 1	55	27133	-	5.0000	7470A	PPB	22 Aug 2023 12:46:00	S	Std
7.5 - 1	57.5	43420	-	7.5000	7470A	PPB	22 Aug 2023 12:48:17	S	Std
10.0 - 1	510	56423	-	10.0000	7470A	PPB	22 Aug 2023 12:50:33	S	Std
ICV27 - 1	ICV27	22522	4.0093	-	7470A	PPB	22 Aug 2023 12:53:39	U	SMPL
ICB27 - 1	ICB27	24	0.0591	-	7470A	PPB	22 Aug 2023 12:55:53	U	SMPL
CCV71 - 1	CCV71	26507	4.7090	-	7470A	PPB	22 Aug 2023 12:58:08	U	SMPL
CCB71 - 1	CCB71	28	0.0598	-	7470A	PPB	22 Aug 2023 13:00:23	U	SMPL
CRA - 1	CRA	984	0.2276	-	7470A	PPB	22 Aug 2023 13:02:38	U	SMPL
HighStd - 1	HighStd	55603	9.8178	-	7470A	PPB	22 Aug 2023 13:04:52	U	SMPL
ChkStd - 1	ChkStd	37163	6.5801	-	7470A	PPB	22 Aug 2023 13:07:07	U	SMPL
PB154982BL - 1	PBW	59	0.0652	-	7470A	PPB	22 Aug 2023 13:09:24	U	SMPL
PB154982BS - 1	LCSW	23076	4.1066	-	7470A	PPB	22 Aug 2023 13:11:40	U	SMPL
O4060-02 - 1	AUD-1038	397	0.1246	-	7470A	PPB	22 Aug 2023 13:13:56	U	SMPL
O4060-02DUP - 1	AUD-1038DUP	430	0.1303	-	7470A	PPB	22 Aug 2023 13:16:12	U	SMPL
O4060-02MS - 1	AUD-1038MS	16133	2.8875	-	7470A	PPB	22 Aug 2023 13:18:29	U	SMPL
O4060-02MSD - 1	AUD-1038MSD	16881	3.0189	-	7470A	PPB	22 Aug 2023 13:20:46	U	SMPL
O4074-01 - 1	MAPLE-SHADE-COMP	7	0.0561	-	7470A	PPB	22 Aug 2023 13:23:02	U	SMPL
CCV72 - 1	CCV72	26688	4.7408	-	7470A	PPB	22 Aug 2023 13:25:17	U	SMPL
CCB72 - 1	CCB72	30	0.0601	-	7470A	PPB	22 Aug 2023 13:27:33	U	SMPL
O4086-21 - 1	FB04	18	0.0580	-	7470A	PPB	22 Aug 2023 13:29:47	U	SMPL
O4093-06 - 1	EB-20230818	23	0.0589	-	7470A	PPB	22 Aug 2023 13:32:03	U	SMPL
PB154983BL - 1	PBW	98	0.0721	-	7470A	PPB	22 Aug 2023 13:34:18	U	SMPL
PB154983BS - 1	LCSW	20671	3.6843	-	7470A	PPB	22 Aug 2023 13:36:34	U	SMPL
O4071-01 - 1	SB-1	53	0.0642	-	7470A	PPB	22 Aug 2023 13:38:50	U	SMPL
O4071-02 - 1	SB-2	108	0.0738	-	7470A	PPB	22 Aug 2023 13:41:06	U	SMPL
O4071-03 - 1	SB-3	110	0.0742	-	7470A	PPB	22 Aug 2023 13:43:22	U	SMPL
O4071-04 - 1	SB-4	177	0.0859	-	7470A	PPB	22 Aug 2023 13:45:38	U	SMPL
O4071-05 - 1	SB-5	112	0.0745	-	7470A	PPB	22 Aug 2023 13:47:55	U	SMPL
O4071-06 - 1	SB-6	271	0.1024	-	7470A	PPB	22 Aug 2023 13:50:10	U	SMPL
CCV73 - 1	CCV73	26569	4.7199	-	7470A	PPB	22 Aug 2023 13:52:25	U	SMPL
CCB73 - 1	CCB73	37	0.0613	-	7470A	PPB	22 Aug 2023 13:54:41	U	SMPL
O4071-07 - 1	SB-7	97	0.0719	-	7470A	PPB	22 Aug 2023 13:56:56	U	SMPL
O4071-08 - 1	SB-8	132	0.0780	-	7470A	PPB	22 Aug 2023 13:59:11	U	SMPL
O4071-09 - 1	SB-9	125	0.0768	-	7470A	PPB	22 Aug 2023 14:01:27	U	SMPL
O4071-10 - 1	SB-10	101	0.0726	-	7470A	PPB	22 Aug 2023 14:03:43	U	SMPL
O4080-02 - 1	ETGI-340	83	0.0694	-	7470A	PPB	22 Aug 2023 14:05:59	U	SMPL
O4082-02 - 1	277	163	0.0835	-	7470A	PPB	22 Aug 2023 14:08:15	U	SMPL
O4082-04 - 1	364	186	0.0875	-	7470A	PPB	22 Aug 2023 14:10:31	U	SMPL
O4082-06 - 1	72-11929	78	0.0685	-	7470A	PPB	22 Aug 2023 14:12:48	U	SMPL
O4082-08 - 1	217	102	0.0728	-	7470A	PPB	22 Aug 2023 14:15:05	U	SMPL
CCV74 - 1	CCV74	26535	4.7140	-	7470A	PPB	22 Aug 2023 14:19:36	U	SMPL
CCB74 - 1	CCB74	25	0.0592	-	7470A	PPB	22 Aug 2023 14:21:51	U	SMPL
O4086-10 - 1	WC-13	145	0.0803	-	7470A	PPB	22 Aug 2023 14:24:06	U	SMPL
O4086-17 - 1	WC-18	141	0.0796	-	7470A	PPB	22 Aug 2023 14:26:21	U	SMPL
O4086-17DUP - 1	WC-18DUP	145	0.0803	-	7470A	PPB	22 Aug 2023 14:28:36	U	SMPL
O4086-17MS - 1	WC-18MS	20345	3.6271	-	7470A	PPB	22 Aug 2023 14:30:56	U	SMPL
O4086-17MSD - 1	WC-18MSD	18971	3.3858	-	7470A	PPB	22 Aug 2023 14:33:12	U	SMPL
PB154930TB - 1	PB154930TB	43	0.0624	-	7470A	PPB	22 Aug 2023 14:35:28	U	SMPL
O4060-02LX5 - 1		103	0.0729	-	7470A	PPB	22 Aug 2023 14:37:44	U	SMPL
O4060-02A - 1		23531	4.1865	-	7470A	PPB	22 Aug 2023 14:40:01	U	SMPL
O4086-17LX5 - 1		7	0.0561	-	7470A	PPB	22 Aug 2023 14:42:17	U	SMPL
O4086-17A - 1		23167	4.1226	-	7470A	PPB	22 Aug 2023 14:44:33	U	SMPL
CCV75 - 1	CCV75	26746	4.7510	-	7470A	PPB	22 Aug 2023 14:46:48	U	SMPL
CCB75 - 1	CCB75	13	0.0571	-	7470A	PPB	22 Aug 2023 14:49:04	U	SMPL