



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB126973

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/23/2023	08:30
2	CAL2	1	Water	NA	NA	20.6	7.01	08/23/2023	08:31
3	CAL3	1	Water	NA	NA	20.7	10.02	08/23/2023	08:33
4	ICV	1	Water	NA	NA	20.6	7.02	08/23/2023	08:35
5	CCV1	1	Water	NA	NA	20.6	2.01	08/23/2023	08:39
6	O4104-01	1	Water	NA	NA	20.4	4.50	08/23/2023	08:40
7	O4121-01	1	Water	NA	NA	20.2	7.43	08/23/2023	08:45
8	O4121-01DUP	1	Water	NA	NA	20.1	7.44	08/23/2023	08:46
9	CCV2	1	Water	NA	NA	20.6	12.02	08/23/2023	08:50

18126473

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 04121

WorkList ID : 172950

Department : Wet-Chemistry

Date : 08-23-2023 09:30:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4104-01 A	POOL-WATER	Water	pH	Cool 4 deg C	STER06	N11	08/07/2023	9040C
O4121-01	IDW-W-20230822	Water	pH	Cool 4 deg C	ECOL03	N11	08/22/2023	9040C

Date/Time 08/23/23 09:35

Raw Sample Received by: JR (WCC)

Raw Sample Relinquished by: CJS

Date/Time 08/23/23 12:10

Raw Sample Received by: CJS

Raw Sample Relinquished by: JR (WCC)