

LB126976


7470A

INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

52821



A= 0.0000e+000

B= 1.8702e-004

C= -1.8477e-002

Rho= 0.9995463

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.005 | 0.005  | 124   | 0.000      | 124   | 0     |       |       |       | -   |
| 0.2    | 0.200  | 0.167 | -0.033 | 992   | 0.0 %      | 992   |       |       |       |       | -17 |
| 2.5    | 2.500  | 2.522 | 0.022  | 13585 | 0.0 %      | 13585 |       |       |       |       | 1   |
| 5.0    | 5.000  | 4.927 | -0.073 | 26446 | 0.0 %      | 26446 |       |       |       |       | -1  |
| 7.5    | 7.500  | 7.718 | 0.218  | 41369 | 0.0 %      | 41369 |       |       |       |       | 3   |
| 10.0   | 10.000 | 9.860 | -0.140 | 52821 | 0.0 %      | 52821 |       |       |       |       | -1  |

## LB126976 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Aug 2023 14:15:01

| Sample ID       | Extended ID | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50          | 124        | -       | 0.0000   | 7470A  | PPB   | 23 Aug 2023 14:25:47 | S    | Std  |
| 0.2 - 1         | 50.2        | 992        | -       | 0.2000   | 7470A  | PPB   | 23 Aug 2023 14:28:04 | S    | Std  |
| 2.5 - 1         | 52.5        | 13585      | -       | 2.5000   | 7470A  | PPB   | 23 Aug 2023 14:30:21 | S    | Std  |
| 5.0 - 1         | 55          | 26446      | -       | 5.0000   | 7470A  | PPB   | 23 Aug 2023 14:32:38 | S    | Std  |
| 7.5 - 1         | 57.5        | 41369      | -       | 7.5000   | 7470A  | PPB   | 23 Aug 2023 14:34:54 | S    | Std  |
| 10.0 - 1        | 510         | 52821      | -       | 10.0000  | 7470A  | PPB   | 23 Aug 2023 14:37:10 | S    | Std  |
| ICV29 - 1       | ICV29       | 21702      | 4.0403  | -        | 7470A  | PPB   | 23 Aug 2023 14:42:40 | U    | SMPL |
| ICB29 - 1       | ICB29       | 63         | -0.0067 | -        | 7470A  | PPB   | 23 Aug 2023 14:44:55 | U    | SMPL |
| CCV83 - 1       | CCV83       | 26096      | 4.8620  | -        | 7470A  | PPB   | 23 Aug 2023 14:47:10 | U    | SMPL |
| CCB83 - 1       | CCB83       | 61         | -0.0071 | -        | 7470A  | PPB   | 23 Aug 2023 14:49:24 | U    | SMPL |
| CRA - 1         | CRA         | 1193       | 0.2046  | -        | 7470A  | PPB   | 23 Aug 2023 14:54:10 | U    | SMPL |
| HighStd - 1     | HighStd     | 52566      | 9.8125  | -        | 7470A  | PPB   | 23 Aug 2023 14:56:24 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 38967      | 7.2692  | -        | 7470A  | PPB   | 23 Aug 2023 14:58:40 | U    | SMPL |
| PB155012BL - 1  | PBW         | 24         | -0.0140 | -        | 7470A  | PPB   | 23 Aug 2023 15:00:55 | U    | SMPL |
| PB155012BS - 1  | LCSW        | 20941      | 3.8979  | -        | 7470A  | PPB   | 23 Aug 2023 15:03:11 | U    | SMPL |
| O4082-10 - 1    | 362         | 27         | -0.0134 | -        | 7470A  | PPB   | 23 Aug 2023 15:05:27 | U    | SMPL |
| O4082-10DUP - 1 | 362DUP      | 75         | -0.0045 | -        | 7470A  | PPB   | 23 Aug 2023 15:07:43 | U    | SMPL |
| O4082-10MS - 1  | 362MS       | 21872      | 4.0720  | -        | 7470A  | PPB   | 23 Aug 2023 15:10:00 | U    | SMPL |
| O4082-10MSD - 1 | 362MSD      | 20615      | 3.8370  | -        | 7470A  | PPB   | 23 Aug 2023 15:12:17 | U    | SMPL |
| O4090-02 - 1    | TP-I        | 42         | -0.0106 | -        | 7470A  | PPB   | 23 Aug 2023 15:14:34 | U    | SMPL |
| CCV84 - 1       | CCV84       | 26896      | 5.0116  | -        | 7470A  | PPB   | 23 Aug 2023 15:16:50 | U    | SMPL |
| CCB84 - 1       | CCB84       | 71         | -0.0052 | -        | 7470A  | PPB   | 23 Aug 2023 15:19:06 | U    | SMPL |
| O4090-05 - 1    | TP-K        | 81         | -0.0033 | -        | 7470A  | PPB   | 23 Aug 2023 15:21:21 | U    | SMPL |
| O4090-08 - 1    | TP-L        | 72         | -0.0050 | -        | 7470A  | PPB   | 23 Aug 2023 15:23:36 | U    | SMPL |
| O4090-11 - 1    | TP-N        | 105        | 0.0012  | -        | 7470A  | PPB   | 23 Aug 2023 15:25:52 | U    | SMPL |
| O4091-02 - 1    | TP-A        | 103        | 0.0008  | -        | 7470A  | PPB   | 23 Aug 2023 15:28:07 | U    | SMPL |
| O4091-05 - 1    | TP-B        | 99         | 0.0000  | -        | 7470A  | PPB   | 23 Aug 2023 15:30:23 | U    | SMPL |
| O4091-08 - 1    | TP-C        | 83         | -0.0030 | -        | 7470A  | PPB   | 23 Aug 2023 15:32:40 | U    | SMPL |
| O4091-11 - 1    | TP-E        | 76         | -0.0043 | -        | 7470A  | PPB   | 23 Aug 2023 15:34:56 | U    | SMPL |
| O4091-14 - 1    | TP-H        | 78         | -0.0039 | -        | 7470A  | PPB   | 23 Aug 2023 15:37:13 | U    | SMPL |
| O4091-17 - 1    | TP-G        | 102        | 0.0006  | -        | 7470A  | PPB   | 23 Aug 2023 15:39:30 | U    | SMPL |
| O4091-20 - 1    | TP-F        | 115        | 0.0030  | -        | 7470A  | PPB   | 23 Aug 2023 15:41:46 | U    | SMPL |
| CCV85 - 1       | CCV85       | 26783      | 4.9905  | -        | 7470A  | PPB   | 23 Aug 2023 15:44:01 | U    | SMPL |
| CCB85 - 1       | CCB85       | 77         | -0.0041 | -        | 7470A  | PPB   | 23 Aug 2023 15:46:17 | U    | SMPL |
| PB154967TB - 1  | PB154967TB  | 278        | 0.0335  | -        | 7470A  | PPB   | 23 Aug 2023 15:54:39 | U    | SMPL |
| O4082-10LX5 - 1 |             | 25         | -0.0138 | -        | 7470A  | PPB   | 23 Aug 2023 15:56:54 | U    | SMPL |
| O4082-10A - 1   |             | 20956      | 3.9007  | -        | 7470A  | PPB   | 23 Aug 2023 15:59:09 | U    | SMPL |
| CCV86 - 1       | CCV86       | 27128      | 5.0550  | -        | 7470A  | PPB   | 23 Aug 2023 16:01:25 | U    | SMPL |
| CCB88 - 1       | CCB86       | 72         | -0.0050 | -        | 7470A  | PPB   | 23 Aug 2023 16:03:40 | U    | SMPL |