

LB126982


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

52567



A= 0.0000e+000

B= 1.8907e-004 slope

C= 1.7963e-002 y-intercept

Rho= 0.9994763

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	108	0.000	108					5
0.05	0.050					504	0				12
0.20	0.200	0.223	0.023	1083	0.0 %	1083					1
2.50	2.500	2.513	0.013	13197	0.0 %	13197					4
5.00	5.000	4.783	-0.217	25204	0.0 %	25204					2
7.50	7.500	7.686	0.186	40556	0.0 %	40556					0
10.0	10.000	9.957	-0.043	52567	0.0 %	52567					

LB126982

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	108	-		0 SFAM01.1	PPB	8/23/2023 16:46	Std
	0.2 S01	1083	-		0.2 SFAM01.1	PPB	8/23/2023 16:48	Std
	2.5 S02	13197	-		2.5 SFAM01.1	PPB	8/23/2023 16:51	Std
	5 S03	25204	-		5 SFAM01.1	PPB	8/23/2023 16:56	Std
	7.5 S04	40556	-		7.5 SFAM01.1	PPB	8/23/2023 17:01	Std
	10 S05	52567	-		10 SFAM01.1	PPB	8/23/2023 17:06	Std
ICV030	ICV030	22087	4.1939 -		SFAM01.1	PPB	8/23/2023 17:08	SMPL
ICB030	ICB030	23	0.0223 -		SFAM01.1	PPB	8/23/2023 17:11	SMPL
CCV086	CCV086	24921	4.7297 -		SFAM01.1	PPB	8/23/2023 17:13	SMPL
CCB086	CCB086	24	0.0225 -		SFAM01.1	PPB	8/23/2023 17:15	SMPL
PB155041BL	PBW041	73	0.0318 -		SFAM01.1	PPB	8/23/2023 17:17	SMPL
O3963-01	MC0AF4	137	0.0439 -		SFAM01.1	PPB	8/23/2023 17:20	SMPL
O3963-02	MC0AF4D	205	0.0567 -		SFAM01.1	PPB	8/23/2023 17:22	SMPL
O3963-03	MC0AF4S	6039	1.1597 -		SFAM01.1	PPB	8/23/2023 17:24	SMPL
PB155042BL	PB042W	36	0.0248 -		SFAM01.1	PPB	8/23/2023 17:26	SMPL
O3964-01	MC0AF5	148	0.0459 -		SFAM01.1	PPB	8/23/2023 17:29	SMPL
O3964-02	MC0AF5D	163	0.0488 -		SFAM01.1	PPB	8/23/2023 17:31	SMPL
O3964-03	MC0AF5S	6087	1.1688 -		SFAM01.1	PPB	8/23/2023 17:36	SMPL
CCV087	CCV087	24300	4.6123 -		SFAM01.1	PPB	8/23/2023 17:38	SMPL
CCB087	CCB087	43	0.0261 -		SFAM01.1	PPB	8/23/2023 17:40	SMPL