



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB126983

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	08/23/2023	16:15
2	CAL2	1	Water	NA	NA	20.7	7.00	08/23/2023	16:16
3	CAL3	1	Water	NA	NA	20.6	10.02	08/23/2023	16:20
4	ICV	1	Water	NA	NA	20.7	7.01	08/23/2023	16:25
5	CCV1	1	Water	NA	NA	20.6	2.01	08/23/2023	16:30
6	O4143-02	1	Water	NA	NA	22.2	6.23	08/23/2023	16:35
7	O4143-02DUP	1	Water	NA	NA	22.3	6.47	08/23/2023	16:36
8	O4143-03	1	Water	NA	NA	22.2	6.48	08/23/2023	16:40
9	CCV2	1	Water	NA	NA	20.6	12.02	08/23/2023	16:42

U3126983

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 04143

WorkList ID : 172976

Department : Wet-Chemistry

Date : 08-23-2023 16:03:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4143-02	50121	Water	pH	Cool 4 deg C	PSEG03	N11	08/23/2023	9040C
O4143-03	50122	Water	pH	Cool 4 deg C	PSEG03	N11	08/23/2023	9040C

Date/Time 08/23/23 16:10

Raw Sample Received by: CP

Raw Sample Relinquished by: CP

Date/Time 08/23/23 17:30

Raw Sample Received by: CP

Raw Sample Relinquished by: CP