


LB126990

SFAM01.1

INSTRUMENT ID: CV1

 $\mu$  Abs.:

53211



Linear

A= 0.0000e+000

B= 1.9025e-004 slope

C= -4.7347e-002 intercept

Rho= 0.9998504

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | 0/1/1 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-------|
| 0.00   | 0.000  | -0.008 | -0.008 | 207   | 0.000      | 207   |       |       |       |       |       |
| 0.05   | 0.050  | 0.048  | -0.002 | 501   | 0.0 %      | 501   |       |       |       |       | -4    |
| 0.20   | 0.200  | 0.193  | -0.007 | 1261  | 0.0 %      | 1261  | 0     |       |       |       | -4    |
| 2.50   | 2.500  | 2.560  | 0.060  | 13707 | 0.0 %      | 13707 |       |       |       |       | 2     |
| 5.00   | 5.000  | 5.018  | 0.018  | 26627 | 0.0 %      | 26627 |       |       |       |       | 0     |
| 7.50   | 7.500  | 7.362  | -0.138 | 38947 | 0.0 %      | 38947 |       |       |       |       | -2    |
| 10.0   | 10.000 | 10.076 | 0.076  | 53211 | 0.0 %      | 53211 |       |       |       |       | 1     |

LB126990  
INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method     | Units | Date            | Type |
|------------|-------------|------------|-----------|----------|------------|-------|-----------------|------|
|            | 0 S0        | 207        | -         |          | 0 SFAM01.1 | PPB   | 8/24/2023 14:53 | Std  |
|            | 0.05 S01    | 501        | -         | 0.05     | SFAM01.1   | PPB   | 8/24/2023 14:55 | Std  |
|            | 0.2 S02     | 1261       | -         | 0.2      | SFAM01.1   | PPB   | 8/24/2023 14:58 | Std  |
|            | 2.5 S03     | 13707      | -         | 2.5      | SFAM01.1   | PPB   | 8/24/2023 15:00 | Std  |
|            | 5 S04       | 26627      | -         | 5        | SFAM01.1   | PPB   | 8/24/2023 15:05 | Std  |
|            | 7.5 S05     | 38947      | -         | 7.5      | SFAM01.1   | PPB   | 8/24/2023 15:07 | Std  |
|            | 10 S06      | 53211      | -         | 10       | SFAM01.1   | PPB   | 8/24/2023 15:15 | Std  |
| ICV031     | ICV031      | 22813      | 4.2929 -  |          | SFAM01.1   | PPB   | 8/24/2023 15:20 | SMPL |
| ICB031     | ICB031      | 125        | -0.0236 - |          | SFAM01.1   | PPB   | 8/24/2023 15:24 | SMPL |
| CCV088     | CCV088      | 26607      | 5.0147 -  |          | SFAM01.1   | PPB   | 8/24/2023 15:27 | SMPL |
| CCB088     | CCB088      | 141        | -0.0205 - |          | SFAM01.1   | PPB   | 8/24/2023 15:31 | SMPL |
| PB155043BL | PBW043      | 125        | -0.0236 - |          | SFAM01.1   | PPB   | 8/24/2023 15:34 | SMPL |
| O4140-01   | EXEP6       | 309        | 0.0114 -  |          | SFAM01.1   | PPB   | 8/24/2023 15:36 | SMPL |
| O4140-02   | EXEP7       | 181        | -0.0129 - |          | SFAM01.1   | PPB   | 8/24/2023 15:38 | SMPL |
| O4140-03   | EXEP8       | 210        | -0.0074 - |          | SFAM01.1   | PPB   | 8/24/2023 15:40 | SMPL |
| O4140-04   | EXEP8D      | 164        | -0.0161 - |          | SFAM01.1   | PPB   | 8/24/2023 15:43 | SMPL |
| O4140-05   | EXEP8S      | 5697       | 1.0365 -  |          | SFAM01.1   | PPB   | 8/24/2023 15:45 | SMPL |
| CCV089     | CCV089      | 26230      | 4.943 -   |          | SFAM01.1   | PPB   | 8/24/2023 15:47 | SMPL |
| CCB089     | CCB089      | 174        | -0.0142 - |          | SFAM01.1   | PPB   | 8/24/2023 16:03 | SMPL |