



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB127002

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/25/2023	07:50
2	CAL2	1	Water	NA	NA	20.7	6.99	08/25/2023	07:51
3	CAL3	1	Water	NA	NA	20.7	10.02	08/25/2023	07:53
4	ICV	1	Water	NA	NA	20.8	7.02	08/25/2023	08:00
5	CCV1	1	Water	NA	NA	20.7	2.01	08/25/2023	08:02
6	O4143-01	1	Solid	20.02	20	21.2	5.09	08/25/2023	08:05
7	O4143-01DUP	1	Solid	20.03	20	21.2	5.10	08/25/2023	08:06
8	CCV2	1	Water	NA	NA	20.6	12.02	08/25/2023	08:10

UR 127002

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 04143-1 WorkList ID : 173008 Department : Wet-Chemistry Date : 08-25-2023 07:39:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4143-01	EOR-0146	Solid	pH	Cool 4 deg C	PSEG03	N11	08/23/2023	9045D

Date/Time 08/25/23 07:40
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: CFS

Date/Time 08/25/23 11:00
Raw Sample Received by: CFS
Raw Sample Relinquished by: JD (WC)