



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB127003

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/25/2023	07:50
2	CAL2	1	Water	NA	NA	20.7	6.99	08/25/2023	07:51
3	CAL3	1	Water	NA	NA	20.7	10.02	08/25/2023	07:53
4	ICV	1	Water	NA	NA	20.7	7.02	08/25/2023	08:00
5	CCV1	1	Water	NA	NA	20.5	2.01	08/25/2023	08:15
6	O4151-01	1	Water	NA	NA	20.4	6.90	08/25/2023	08:20
7	O4151-01DUP	1	Water	NA	NA	20.5	6.91	08/25/2023	08:21
8	O4152-01	1	Water	NA	NA	20.3	6.51	08/25/2023	08:25
9	CCV2	1	Water	NA	NA	20.6	12.02	08/25/2023	08:30

VB127003

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 04151

WorkList ID : 173007

Department : Wet-Chemistry

Date : 08-25-2023 07:38:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4151-01	EQ-TANK-DNMPL	Water	pH	Cool 4 deg C	PSEG03	N11	08/24/2023	9040C
O4152-01	COMPOSITE-1	Water	pH	Cool 4 deg C	SOUN03	N11	08/23/2023	9040C

Date/Time 08/25/23 07:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 08/25/23 12:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]