



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Sohil

Run Number: LB127004

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/25/2023	07:50
2	CAL2	1	Water	NA	NA	20.7	6.99	08/25/2023	07:51
3	CAL3	1	Water	NA	NA	20.7	10.02	08/25/2023	07:53
4	ICV	1	Water	NA	NA	20.8	7.02	08/25/2023	08:00
5	CCV1	1	Water	NA	NA	20.6	2.02	08/25/2023	08:35
6	O4168-01	1	Solid	20.00	20	20.9	5.06	08/25/2023	08:40
7	O4168-02	1	Solid	20.03	20	20.1	7.48	08/25/2023	08:45
8	O4168-02DUP	1	Solid	20.04	20	20.2	7.49	08/25/2023	08:46
9	CCV2	1	Water	NA	NA	20.7	12.02	08/25/2023	08:50

WORKLIST(Hardcopy Internal Chain)

B127004

WorkList Name : corrosivity o5174

WorkList ID : 173009

Department : Wet-Chemistry

Date : 08-25-2023 07:40:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4168-01	210	Solid	Corrosivity	Cool 4 deg C	PSEG03	N11	08/24/2023	9045D
O4168-02	205	Solid	Corrosivity	Cool 4 deg C	PSEG03	N11	08/24/2023	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/25/23 07:14

JP/CDC

JP/CDC

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/25/23

JP/CDC

JP/CDC

12:00

JP/CDC

JP/CDC