



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127005

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/25/2023	07:50
2	CAL2	1	Water	NA	NA	20.7	6.99	08/25/2023	07:51
3	CAL3	1	Water	NA	NA	20.7	10.02	08/25/2023	07:53
4	ICV	1	Water	NA	NA	20.8	7.02	08/25/2023	08:00
5	CCV1	1	Water	NA	NA	20.5	2.02	08/25/2023	08:50
6	O4165-05	1	Water	NA	NA	20.1	7.63	08/25/2023	09:00
7	O4165-05DUP	1	Water	NA	NA	20.2	7.64	08/25/2023	09:01
8	O4168-03	1	Water	NA	NA	20.4	7.22	08/25/2023	09:05
9	CCV2	1	Water	NA	NA	20.6	12.02	08/25/2023	09:10

VB 127005

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o5165 WorkList ID : 173010 Department : Wet-Chemistry Date : 08-25-2023 07:41:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4165-05	2909	Water	Corrosivity	Cool 4 deg C	PSEG03	N11	08/24/2023	9040C
O4168-03	209	Water	Corrosivity	Cool 4 deg C	PSEG03	N11	08/24/2023	9040C

Date/Time 08/25/23 07:45
Raw Sample Received by: JPC
Raw Sample Relinquished by: JPC

Date/Time 08/25/23 12:00
Raw Sample Received by: JPC
Raw Sample Relinquished by: JPC