


LB127006

7470A

INSTRUMENT ID: CV1

 μ Abs.:

55535



Linear

A= 0.0000e+000

B= 1.8390e-004

C= -1.4891e-001

Rho= 0.9999058

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.008	0.008	855	0.000	855					
0.2	0.200	0.196	-0.004	1875	0.0 %	1875	0				-2
2.5	2.500	2.523	0.023	14530	0.0 %	14530					1
5.0	5.000	5.011	0.011	28057	0.0 %	28057					0
7.5	7.500	7.397	-0.103	41034	0.0 %	41034					-1
10.0	10.000	10.064	0.064	55535	0.0 %	55535					1

LB127006 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Aug 2023 10:34:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	855	-	0.0000	7470A	PPB	25 Aug 2023 10:44:10	S	Std
0.2 - 1	50.2	1875	-	0.2000	7470A	PPB	25 Aug 2023 10:46:27	S	Std
2.5 - 1	52.5	14530	-	2.5000	7470A	PPB	25 Aug 2023 10:48:44	S	Std
5.0 - 1	55	28057	-	5.0000	7470A	PPB	25 Aug 2023 10:51:01	S	Std
7.5 - 1	57.5	41034	-	7.5000	7470A	PPB	25 Aug 2023 10:53:18	S	Std
10.0 - 1	510	55535	-	10.0000	7470A	PPB	25 Aug 2023 10:58:07	S	Std
ICV32 - 1	ICV32	23416	4.1574	-	7470A	PPB	25 Aug 2023 11:02:56	U	SMPL
ICB32 - 1	ICB32	798	-0.0022	-	7470A	PPB	25 Aug 2023 11:05:11	U	SMPL
CCV90 - 1	CCV90	27793	4.9623	-	7470A	PPB	25 Aug 2023 11:07:26	U	SMPL
CCB90 - 1	CCB90	808	-0.0003	-	7470A	PPB	25 Aug 2023 11:09:40	U	SMPL
CRA - 1	CRA	1843	0.1900	-	7470A	PPB	25 Aug 2023 11:11:55	U	SMPL
HighStd - 1	HighStd	54881	9.9439	-	7470A	PPB	25 Aug 2023 11:14:10	U	SMPL
ChkStd - 1	ChkStd	35787	6.4325	-	7470A	PPB	25 Aug 2023 11:16:25	U	SMPL
PB155058BL - 1	PBW	852	0.0078	-	7470A	PPB	25 Aug 2023 11:18:41	U	SMPL
PB155058BS - 1	LCSW	19666	3.4678	-	7470A	PPB	25 Aug 2023 11:20:57	U	SMPL
O4122-01 - 1	FLY-ASH	1008	0.0365	-	7470A	PPB	25 Aug 2023 11:23:12	U	SMPL
O4122-01DUP - 1	FLY-ASHDUP	1061	0.0462	-	7470A	PPB	25 Aug 2023 11:25:28	U	SMPL
O4122-01MS - 1	FLY-ASHMS	20218	3.5693	-	7470A	PPB	25 Aug 2023 11:27:44	U	SMPL
O4122-01MSD - 1	FLY-ASHMSD	19925	3.5154	-	7470A	PPB	25 Aug 2023 11:30:01	U	SMPL
PB155059BL - 1	PBW	776	-0.0062	-	7470A	PPB	25 Aug 2023 11:32:16	U	SMPL
CCV91 - 1	CCV91	28124	5.0232	-	7470A	PPB	25 Aug 2023 11:34:32	U	SMPL
CCB91 - 1	CCB91	825	0.0028	-	7470A	PPB	25 Aug 2023 11:36:46	U	SMPL
PB155059BS - 1	LCSW	19505	3.4381	-	7470A	PPB	25 Aug 2023 11:39:02	U	SMPL
O4121-01 - 1	IDW-W-20230822	1054	0.0449	-	7470A	PPB	25 Aug 2023 11:41:16	U	SMPL
O4121-01DUP - 1	IDW-W-20230822DUP	459	-0.0645	-	7470A	PPB	25 Aug 2023 11:43:31	U	SMPL
O4121-01MS - 1	IDW-W-20230822MS	20630	3.6450	-	7470A	PPB	25 Aug 2023 11:45:47	U	SMPL
O4121-01MSD - 1	IDW-W-20230822MSD	19702	3.4744	-	7470A	PPB	25 Aug 2023 11:48:03	U	SMPL
PB155060BL - 1	PBW	911	0.0186	-	7470A	PPB	25 Aug 2023 11:50:19	U	SMPL
PB155060BS - 1	LCSW	19988	3.5270	-	7470A	PPB	25 Aug 2023 11:55:24	U	SMPL
O4109-07 - 1	RINSATE-BLANK	740	-0.0128	-	7470A	PPB	25 Aug 2023 11:57:40	U	SMPL
O4110-21 - 1	RINSATE-BLANK	815	0.0010	-	7470A	PPB	25 Aug 2023 11:59:57	U	SMPL
CCV92 - 1	CCV92	27886	4.9794	-	7470A	PPB	25 Aug 2023 12:02:12	U	SMPL
CCB92 - 1	CCB92	828	0.0034	-	7470A	PPB	25 Aug 2023 12:04:28	U	SMPL
PB155045TB - 1	PB155045TB	919	0.0201	-	7470A	PPB	25 Aug 2023 12:06:43	U	SMPL
PB155024TB - 1	PB155024TB	895	0.0157	-	7470A	PPB	25 Aug 2023 12:08:59	U	SMPL
O4122-01LX5 - 1		180	-0.1158	-	7470A	PPB	25 Aug 2023 12:11:14	U	SMPL
O4122-01A - 1		23558	4.1835	-	7470A	PPB	25 Aug 2023 12:13:29	U	SMPL
O4121-01LX5 - 1		201	-0.1119	-	7470A	PPB	25 Aug 2023 12:15:44	U	SMPL
O4121-01A - 1		22986	4.0783	-	7470A	PPB	25 Aug 2023 12:18:00	U	SMPL
CCV93 - 1	CCV93	27475	4.9039	-	7470A	PPB	25 Aug 2023 12:20:16	U	SMPL
CCB93 - 1	CCB93	810	0.0001	-	7470A	PPB	25 Aug 2023 12:22:32	U	SMPL