

LB127017


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52068



A= 0.0000e+000

B= 1.9408e-004

C= -1.6458e-001

Rho= 0.9999517

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/1
0.0	0.000	-0.003	-0.003	835	0.000	835					1
0.05	0.050	0.043	-0.007	1072	0.0 %	1072	0				14
0.2	0.200	0.195	-0.005	1855	0.0 %	1855					3
2.5	2.500	2.483	-0.017	13641	0.0 %	13641					1
5.0	5.000	5.019	0.019	26709	0.0 %	26709					0
7.5	7.500	7.571	0.071	39859	0.0 %	39859					1
10.0	10.000	9.941	-0.059	52068	0.0 %	52068					1

LB127017 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Aug 2023 16:13:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	835	-	0.0000	245.1	PPB	25 Aug 2023 16:15:47	S	Std
0.05 - 1	50.05	1072	-	0.0500	245.1	PPB	25 Aug 2023 16:18:04	S	Std
0.2 - 1	50.2	1855	-	0.2000	245.1	PPB	25 Aug 2023 16:20:21	S	Std
2.5 - 1	52.5	13641	-	2.5000	245.1	PPB	25 Aug 2023 16:22:37	S	Std
5.0 - 1	55	26709	-	5.0000	245.1	PPB	25 Aug 2023 16:24:53	S	Std
7.5 - 1	57.5	39859	-	7.5000	245.1	PPB	25 Aug 2023 16:27:09	S	Std
10.0 - 1	510	52068	-	10.0000	245.1	PPB	25 Aug 2023 16:48:20	S	Std
ICV34 - 1	ICV34	21122	3.9347	-245.1		PPB	25 Aug 2023 16:51:20	U	SMPL
ICB34 - 1	ICB34	724	-0.0241	-245.1		PPB	25 Aug 2023 16:53:35	U	SMPL
CCV01 - 1	CCV01	26908	5.0576	-245.1		PPB	25 Aug 2023 16:55:49	U	SMPL
CCB01 - 1	CCB01	720	-0.0248	-245.1		PPB	25 Aug 2023 16:58:04	U	SMPL
CRA - 1	CRA	1083	0.0456	-245.1		PPB	25 Aug 2023 17:00:19	U	SMPL
HighStd - 1	HighStd	52107	9.9482	-245.1		PPB	25 Aug 2023 17:07:27	U	SMPL
ChkStd - 1	ChkStd	38271	7.2629	-245.1		PPB	25 Aug 2023 17:09:41	U	SMPL
PB155107BL - 1	PBW	871	0.0045	-245.1		PPB	25 Aug 2023 17:11:57	U	SMPL
PB155107BS - 1	LCSW	21234	3.9565	-245.1		PPB	25 Aug 2023 17:14:12	U	SMPL
O4152-01 - 1	COMPOSITE-1	36678	6.9538	-245.1		PPB	25 Aug 2023 17:16:28	U	SMPL
O4152-01DUP - 1	COMPOSITE-1DUP	35099	6.6473	-245.1		PPB	25 Aug 2023 17:18:45	U	SMPL
O4152-01MS - 1	COMPOSITE-1MS	52029	9.9331	-245.1		PPB	25 Aug 2023 17:23:42	U	SMPL
O4152-01MSD - 1	COMPOSITE-1MSD	52154	9.9573	-245.1		PPB	25 Aug 2023 17:25:58	U	SMPL
CCV02 - 1	CCV02	26798	5.0363	-245.1		PPB	25 Aug 2023 17:28:14	U	SMPL
CCB02 - 1	CCB02	707	-0.0274	-245.1		PPB	25 Aug 2023 17:30:29	U	SMPL
O4152-01LX5 - 1		6508	1.0985	-245.1		PPB	25 Aug 2023 17:32:44	U	SMPL
O4152-01A - 1		47595	9.0725	-245.1		PPB	25 Aug 2023 17:37:50	U	SMPL
CCV03 - 1	CCV03	26985	5.0726	-245.1		PPB	25 Aug 2023 17:40:05	U	SMPL
CCB03 - 1	CCB03	780	-0.0132	-245.1		PPB	25 Aug 2023 17:42:20	U	SMPL