



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/28/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 18:10
In Date: 08/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:26
Out Date: 08/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127018

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4114-01	MEZYH4	1	1.13	8.55	9.68	9.04	92.5	
O4114-02	MEZYH5	2	1.13	8.37	9.5	8.45	87.5	
O4114-03	MEZYH5D	3	1.13	8.37	9.5	8.45	87.5	
O4114-04	MEZYH5S	4	1.13	8.37	9.5	8.45	87.5	
O4114-05	MEZYH6	5	1.1	8.79	9.89	9.13	91.4	
O4114-06	MEZYL1	6	1.18	8.73	9.91	9.16	91.4	
O4114-07	MEZYL2	7	1.15	8.62	9.77	8.68	87.4	
O4114-08	MEZYL3	8	1.15	8.84	9.99	9.18	90.8	
O4114-09	MEZYH7	9	1.18	8.54	9.72	8.89	90.3	
O4114-10	MEZYH8	10	1.18	8.56	9.74	8.8	89.0	
O4114-11	MEZYH9	11	1.17	8.46	9.63	8.59	87.7	
O4114-13	MEZYJ0	12	1.19	8.63	9.82	8.65	86.4	
O4114-14	MEZYJ1	13	1.16	8.50	9.66	8.24	83.3	
O4114-15	MEZYJ2	14	1.14	8.84	9.98	8.45	82.7	
O4114-16	MEZYG5	15	1.15	8.38	9.53	8.77	90.9	
O4114-17	MEZYG6	16	1.18	8.75	9.93	9.03	89.7	
O4114-18	MEZYG7	17	1.19	8.73	9.92	8.49	83.6	
O4114-19	MEZYG7	17	1.19	8.73	9.92	8.49	83.6	
O4114-19	MEZYG7	17	1.19	8.73	9.92	8.49	83.6	
O4114-19	MEZYX0	18	1.17	8.55	9.72	8.8	89.2	
O4114-20	MEZYX3	19	1.19	8.35	9.54	8.53	87.9	
O4114-21	MEZYX4	20	1.11	8.56	9.67	7.95	79.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$