



PERCENT SOLID

Supervisor: Sohil  
Analyst: jignesh  
Date: 8/28/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 18:10  
In Date: 08/25/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:26  
Out Date: 08/26/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB127018

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O4114-01 | MEZYH4          | 1      | 1.13           | 8.55         | 9.68                    | 9.04                      | 92.5    |          |
| O4114-02 | MEZYH5          | 2      | 1.13           | 8.37         | 9.5                     | 8.45                      | 87.5    |          |
| O4114-03 | MEZYH5D         | 3      | 1.13           | 8.37         | 9.5                     | 8.45                      | 87.5    |          |
| O4114-04 | MEZYH5S         | 4      | 1.13           | 8.37         | 9.5                     | 8.45                      | 87.5    |          |
| O4114-05 | MEZYH6          | 5      | 1.1            | 8.79         | 9.89                    | 9.13                      | 91.4    |          |
| O4114-06 | MEZYL1          | 6      | 1.18           | 8.73         | 9.91                    | 9.16                      | 91.4    |          |
| O4114-07 | MEZYL2          | 7      | 1.15           | 8.62         | 9.77                    | 8.68                      | 87.4    |          |
| O4114-08 | MEZYL3          | 8      | 1.15           | 8.84         | 9.99                    | 9.18                      | 90.8    |          |
| O4114-09 | MEZYH7          | 9      | 1.18           | 8.54         | 9.72                    | 8.89                      | 90.3    |          |
| O4114-10 | MEZYH8          | 10     | 1.18           | 8.56         | 9.74                    | 8.8                       | 89.0    |          |
| O4114-11 | MEZYH9          | 11     | 1.17           | 8.46         | 9.63                    | 8.59                      | 87.7    |          |
| O4114-13 | MEZYG0          | 12     | 1.19           | 8.63         | 9.82                    | 8.65                      | 86.4    |          |
| O4114-14 | MEZYG1          | 13     | 1.16           | 8.50         | 9.66                    | 8.24                      | 83.3    |          |
| O4114-15 | MEZYG2          | 14     | 1.14           | 8.84         | 9.98                    | 8.45                      | 82.7    |          |
| O4114-16 | MEZYG5          | 15     | 1.15           | 8.38         | 9.53                    | 8.77                      | 90.9    |          |
| O4114-17 | MEZYG6          | 16     | 1.18           | 8.75         | 9.93                    | 9.03                      | 89.7    |          |
| O4114-18 | MEZYG7          | 17     | 1.19           | 8.73         | 9.92                    | 8.49                      | 83.6    |          |
| O4114-19 | MEZYGX0         | 18     | 1.17           | 8.55         | 9.72                    | 8.8                       | 89.2    |          |
| O4114-20 | MEZYGX3         | 19     | 1.19           | 8.35         | 9.54                    | 8.53                      | 87.9    |          |
| O4114-21 | MEZYGX4         | 20     | 1.11           | 8.56         | 9.67                    | 7.95                      | 79.9    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

WP 127018

WorkList Name : %1-O4114

WorkList ID : 173056

Department : Wet-Chemistry

Date : 08-25-2023 17:10:54

| Sample   | Customer Sample | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-----------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| O4114-01 | MEZYH4          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-02 | MEZYH5          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-03 | MEZYH5D         | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-04 | MEZYH5S         | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-05 | MEZYH6          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-06 | MEZYL1          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-07 | MEZYL2          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-08 | MEZYL3          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-09 | MEZYH7          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/21/2023   | Chemtech -SO |
| O4114-10 | MEZYH8          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/22/2023   | Chemtech -SO |
| O4114-11 | MEZYH9          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/22/2023   | Chemtech -SO |
| O4114-13 | MEZYJ0          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/22/2023   | Chemtech -SO |
| O4114-14 | MEZYJ1          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/23/2023   | Chemtech -SO |
| O4114-15 | MEZYJ2          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/23/2023   | Chemtech -SO |
| O4114-16 | MEZYG5          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/23/2023   | Chemtech -SO |
| O4114-17 | MEZYG6          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/24/2023   | Chemtech -SO |
| O4114-18 | MEZYG7          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/24/2023   | Chemtech -SO |
| O4114-19 | MEZYX0          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/24/2023   | Chemtech -SO |
| O4114-20 | MEZYX3          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/24/2023   | Chemtech -SO |
| O4114-21 | MEZYX4          | Solid  | Percent Solids | Cool 4 deg C | USEP01   | A53                         | 08/24/2023   | Chemtech -SO |

Date/Time 08/25/23 17:15  
 Raw Sample Received by: TP (wpc)  
 Raw Sample Relinquished by: DP (SM)

Date/Time 08/25/23 18:20  
 Raw Sample Received by: DP (SM)  
 Raw Sample Relinquished by: DP (wpc)