



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Sohil

Run Number: LB127028

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/26/2023	10:40
2	O4143-02		1	12	100.00	>212.0	>212.0	08/26/2023	11:10
3	O4143-02DUP		1	13	100.00	>212.0	>212.0	08/26/2023	11:40
4	O4143-03		1	12	100.00	>212.0	>212.0	08/26/2023	12:10
5	O4152-01		1	13	100.00	>212.0	>212.0	08/26/2023	12:40
6	O4165-05		1	14	100.00	>212.0	>212.0	08/26/2023	13:10
7	O4168-03		1	14	100.00	>212.0	>212.0	08/26/2023	13:40
8	O4171-01		1	12	100.00	>212.0	>212.0	08/26/2023	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26127028

WorkList Name : fp-8-22

WorkList ID : 172912

Department : Wet-Chemistry

Date : 08-22-2023 08:06:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4143-02	50121	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/23/2023	1010B
O4143-03	50122	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/23/2023	1010B
O4165-05	2909	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/24/2023	1010B
O4168-03	209	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/24/2023	1010B
O4171-01	FMI-134	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/25/2023	1010B
O4152-01	COMPOSITE-1	Water	Flash Point	Cool 4 deg C	SOUN03		08/23/2023	1010B

Date/Time 08/26/2023 10:30
 Raw Sample Received by: RFL
 Raw Sample Relinquished by: JH6001

Date/Time 08/26/2023 15:20
 Raw Sample Received by: JH6001
 Raw Sample Relinquished by: RFL