

LB127044


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

60992



A= 0.0000e+000

B= 1.6443e-004

C= -2.8730e-002

Rho= 0.9999298

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.065	0.065	568	0.000	568				
0.2	0.200	0.189	-0.011	1322	0.0 %	1322				
2.5	2.500	2.418	-0.082	14881	0.0 %	14881				
5.0	5.000	5.008	0.008	30629	0.0 %	30629				
7.5	7.500	7.520	0.020	45909	0.0 %	45909				
10.0	10.000	10.000	0.000	60992	0.0 %	60992				

0.0000
1536
0.0000

LB127044 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Aug 2023 16:22:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	568	-	0.0000	7471B	PPB	28 Aug 2023 16:23:59	S	Std
0.2 - 1	50.2	1322	-	0.2000	7471B	PPB	28 Aug 2023 16:26:15	S	Std
2.5 - 1	52.5	14881	-	2.5000	7471B	PPB	28 Aug 2023 16:28:32	S	Std
5.0 - 1	55	30629	-	5.0000	7471B	PPB	28 Aug 2023 16:30:49	S	Std
7.5 - 1	57.5	45909	-	7.5000	7471B	PPB	28 Aug 2023 16:33:05	S	Std
10.0 - 1	510	60992	-	10.0000	7471B	PPB	28 Aug 2023 16:35:22	S	Std
ICV36 - 1	ICV36	25663	4.1911	-	7471B	PPB	28 Aug 2023 16:38:10	U	SMPL
ICB36 - 1	ICB36	455	0.0461	-	7471B	PPB	28 Aug 2023 16:40:25	U	SMPL
CCV09 - 1	CCV09	30698	5.0191	-	7471B	PPB	28 Aug 2023 16:42:41	U	SMPL
CCB09 - 1	CCB09	441	0.0438	-	7471B	PPB	28 Aug 2023 16:44:55	U	SMPL
CRA - 1	CRA	1299	0.1849	-	7471B	PPB	28 Aug 2023 16:47:11	U	SMPL
HighStd - 1	HighStd	61206	10.0356	-	7471B	PPB	28 Aug 2023 16:49:27	U	SMPL
ChkStd - 1	ChkStd	41783	6.8418	-	7471B	PPB	28 Aug 2023 16:51:42	U	SMPL
PB155163BL - 1	PBS	206	0.0051	-	7471B	PPB	28 Aug 2023 16:53:59	U	SMPL
PB155163BS - 1	LCSS	22972	3.7487	-	7471B	PPB	28 Aug 2023 16:58:53	U	SMPL
O4170-01 - 1	WC-SCAS-01	3967	0.6236	-	7471B	PPB	28 Aug 2023 17:01:08	U	SMPL
O4177-01 - 1	EO-1-082523	1666	0.2452	-	7471B	PPB	28 Aug 2023 17:03:24	U	SMPL
CCV10 - 1	CCV10	30350	4.9619	-	7471B	PPB	28 Aug 2023 17:12:29	U	SMPL
CCB10 - 1	CCB10	446	0.0446	-	7471B	PPB	28 Aug 2023 17:14:45	U	SMPL
O4183-03 - 1	HR-03-082523	5708	0.9099	-	7471B	PPB	28 Aug 2023 17:19:17	U	SMPL
O4183-05 - 1	HR-04-082523	1487	0.2158	-	7471B	PPB	28 Aug 2023 17:21:33	U	SMPL
O4188-01 - 1	WC-7	3666	0.5741	-	7471B	PPB	28 Aug 2023 17:23:48	U	SMPL
O4188-04 - 1	WC-6	2293	0.3483	-	7471B	PPB	28 Aug 2023 17:26:03	U	SMPL
O4188-07 - 1	WC-5	923	0.1230	-	7471B	PPB	28 Aug 2023 17:28:19	U	SMPL
O4188-10 - 1	WC-4	7241	1.1619	-	7471B	PPB	28 Aug 2023 17:30:36	U	SMPL
O4189-01 - 1	SB-02(22-23)	292333	48.0408	-	7471B	PPB	28 Aug 2023 17:32:53	U	SMPL
O4189-02 - 1	SB-06(16-18)	-40	-0.0353	-	7471B	PPB	28 Aug 2023 17:35:09	U	SMPL
O4189-03 - 1	SB-07(14-15)	995	0.1349	-	7471B	PPB	28 Aug 2023 17:37:41	U	SMPL
CCV11 - 1	CCV11	30524	4.9905	-	7471B	PPB	28 Aug 2023 17:39:56	U	SMPL
CCB11 - 1	CCB11	425	0.0412	-	7471B	PPB	28 Aug 2023 17:42:12	U	SMPL