

LB127048


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

61673



A= 0.0000e+000

B= 1.5846e-004

C= -1.4341e-002

Rho= 0.9989911

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/60
0.0	0.000	0.013	0.013	173	0.000	173					1
0.2	0.200	0.181	-0.019	1234	0.0 %	1234					10
2.5	2.500	2.398	-0.102	15222	0.0 %	15222					4
5.0	5.000	5.044	0.044	31925	0.0 %	31925					1
7.5	7.500	7.805	0.305	49350	0.0 %	49350					4
10.0	10.000	9.758	-0.242	61673	0.0 %	61673					2

LB127048 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Aug 2023 10:45:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	173	-	0.0000	7471B	PPB	29 Aug 2023 10:48:30	S	Std
0.2 - 1	50.2	1234	-	0.2000	7471B	PPB	29 Aug 2023 10:50:47	S	Std
2.5 - 1	52.5	15222	-	2.5000	7471B	PPB	29 Aug 2023 10:53:03	S	Std
5.0 - 1	55	31925	-	5.0000	7471B	PPB	29 Aug 2023 10:55:20	S	Std
7.5 - 1	57.5	49350	-	7.5000	7471B	PPB	29 Aug 2023 10:57:37	S	Std
10.0 - 1	510	61673	-	10.0000	7471B	PPB	29 Aug 2023 10:59:52	S	Std
ICV37 - 1	ICV37	26400	4.1689	-	7471B	PPB	29 Aug 2023 11:02:44	U	SMPL
ICB37 - 1	ICB37	93	0.0004	-	7471B	PPB	29 Aug 2023 11:04:59	U	SMPL
CCV13 - 1	CCV13	31618	4.9957	-	7471B	PPB	29 Aug 2023 11:07:14	U	SMPL
CCB13 - 1	CCB13	100	0.0015	-	7471B	PPB	29 Aug 2023 11:09:29	U	SMPL
CRA - 1	CRA	1202	0.1761	-	7471B	PPB	29 Aug 2023 11:11:44	U	SMPL
HighStd - 1	HighStd	61540	9.7371	-	7471B	PPB	29 Aug 2023 11:13:59	U	SMPL
ChkStd - 1	ChkStd	44390	7.0196	-	7471B	PPB	29 Aug 2023 11:16:14	U	SMPL
PB155177BL - 1	PBS	23	-0.0107	-	7471B	PPB	29 Aug 2023 11:18:31	U	SMPL
PB155177BS - 1	LCSS	24182	3.8175	-	7471B	PPB	29 Aug 2023 11:20:48	U	SMPL
O4196-01 - 1	OR-34-COMP-01	11896	1.8707	-	7471B	PPB	29 Aug 2023 11:23:03	U	SMPL
O4196-01DUP - 1	OR-34-COMP-01DUP	12825	2.0179	-	7471B	PPB	29 Aug 2023 11:25:19	U	SMPL
O4196-01MS - 1	OR-34-COMP-01MS	40763	6.4448	-	7471B	PPB	29 Aug 2023 11:27:36	U	SMPL
O4196-01MSD - 1	OR-34-COMP-01MSD	40789	6.4489	-	7471B	PPB	29 Aug 2023 11:31:58	U	SMPL
O4196-08 - 1	OR-345-COMP-02	2364	0.3603	-	7471B	PPB	29 Aug 2023 11:34:14	U	SMPL
CCV14 - 1	CCV14	31105	4.9145	-	7471B	PPB	29 Aug 2023 11:36:30	U	SMPL
CCB14 - 1	CCB14	105	0.0023	-	7471B	PPB	29 Aug 2023 11:38:45	U	SMPL
O4196-15 - 1	OR-345-COMP-03	1674	0.2509	-	7471B	PPB	29 Aug 2023 11:41:00	U	SMPL
O4196-22 - 1	OR-345-COMP-04	849	0.1202	-	7471B	PPB	29 Aug 2023 11:43:15	U	SMPL
O4196-29 - 1	OR-345-COMP-05	2712	0.4154	-	7471B	PPB	29 Aug 2023 11:45:31	U	SMPL
O4197-01 - 1	OR-309-OK-370-COMP-20	1247	0.1833	-	7471B	PPB	29 Aug 2023 11:47:46	U	SMPL
O4197-08 - 1	OR-309-OK-370-COMP-21	2263	0.3442	-	7471B	PPB	29 Aug 2023 11:50:01	U	SMPL
O4197-15 - 1	OR-309-OK-370-COMP-22	1787	0.2688	-	7471B	PPB	29 Aug 2023 11:52:17	U	SMPL
O4197-22 - 1	OR-309-OK-370-COMP-23	1151	0.1680	-	7471B	PPB	29 Aug 2023 11:54:33	U	SMPL
O4198-01 - 1	OR-309-OK-370-COMP-24	797	0.1119	-	7471B	PPB	29 Aug 2023 11:56:50	U	SMPL
O4198-07 - 1	OR-309-OK-370-COMP-25	1268	0.1866	-	7471B	PPB	29 Aug 2023 11:59:07	U	SMPL
O4198-14 - 1	OR-309-OK-370-COMP-26	820	0.1156	-	7471B	PPB	29 Aug 2023 12:01:23	U	SMPL
CCV15 - 1	CCV15	31241	4.9360	-	7471B	PPB	29 Aug 2023 12:03:38	U	SMPL
CCB15 - 1	CCB15	108	0.0028	-	7471B	PPB	29 Aug 2023 12:05:53	U	SMPL
O4202-01 - 1	C-1	71754	11.3556	-	7471B	PPB	29 Aug 2023 12:08:09	U	SMPL
O4202-02 - 1	C-2	3053	0.4694	-	7471B	PPB	29 Aug 2023 12:10:24	U	SMPL
O4202-03 - 1	C-3	28286	4.4678	-	7471B	PPB	29 Aug 2023 12:12:41	U	SMPL
O4183-01 - 1	HR-02-082523	3549	0.5480	-	7471B	PPB	29 Aug 2023 12:14:56	U	SMPL
O4183-01DUP - 1	HR-02-082523DUP	3522	0.5437	-	7471B	PPB	29 Aug 2023 12:17:13	U	SMPL
O4183-01MS - 1	HR-02-082523MS	25263	3.9888	-	7471B	PPB	29 Aug 2023 12:19:29	U	SMPL
O4183-01MSD - 1	HR-02-082523MSD	25179	3.9754	-	7471B	PPB	29 Aug 2023 12:21:45	U	SMPL
O4183-01LX5 - 1		674	0.0925	-	7471B	PPB	29 Aug 2023 12:24:01	U	SMPL
O4183-01A - 1		26215	4.1396	-	7471B	PPB	29 Aug 2023 12:26:18	U	SMPL
O4189-04 - 1	SB-01(48-50)	348	0.0408	-	7471B	PPB	29 Aug 2023 12:28:34	U	SMPL
CCV16 - 1	CCV16	31239	4.9357	-	7471B	PPB	29 Aug 2023 12:30:49	U	SMPL
CCB16 - 1	CCB16	105	0.0023	-	7471B	PPB	29 Aug 2023 12:33:04	U	SMPL
O4196-01LX5 - 1		2421	0.3693	-	7471B	PPB	29 Aug 2023 12:35:19	U	SMPL
O4196-01A - 1		37008	5.8498	-	7471B	PPB	29 Aug 2023 12:37:35	U	SMPL
O4189-01DLX10 - 1		33077	5.2269	-	7471B	PPB	29 Aug 2023 12:42:38	U	SMPL
O4202-01DLX2 - 1		36688	5.7991	-	7471B	PPB	29 Aug 2023 12:44:53	U	SMPL
CCV17 - 1	CCV17	31333	4.9506	-	7471B	PPB	29 Aug 2023 12:47:09	U	SMPL
CCB17 - 1	CCB17	90	-0.0001	-	7471B	PPB	29 Aug 2023 12:49:25	U	SMPL