



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB127051

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/29/2023	11:35
2	CAL2	1	Water	NA	NA	20.7	7.00	08/29/2023	11:36
3	CAL3	1	Water	NA	NA	20.7	10.03	08/29/2023	11:37
4	ICV	1	Water	NA	NA	20.6	7.03	08/29/2023	11:40
5	CCV1	1	Water	NA	NA	20.6	2.01	08/29/2023	11:44
6	O4171-01	1	Water	NA	NA	20.2	7.54	08/29/2023	11:45
7	O4171-01DUP	1	Water	NA	NA	20.3	7.55	08/29/2023	11:46
8	O4216-01	1	Water	NA	NA	20.1	4.40	08/29/2023	11:50
9	O4216-02	1	Water	NA	NA	20.9	4.42	08/29/2023	11:51
10	CCV2	1	Water	NA	NA	20.7	12.02	08/29/2023	12:00

13127051

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 04215 WorkList ID : 173125 Department : Wet-Chemistry Date : 08-29-2023 11:25:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4171-01	FMI-134	Water	pH	Cool 4 deg C	PSEG03	N11	08/25/2023	9040C
O4216-01	MINI-CELL	Water	pH	Cool 4 deg C	STER14	L11	08/25/2023	9040C
O4216-02	JS10000	Water	pH	Cool 4 deg C	STER14	L11	08/25/2023	9040C

Date/Time 08.29.27 11:30

Raw Sample Received by: JS10000

Raw Sample Relinquished by: JS10000

Date/Time 08.29.23 11:00

Raw Sample Received by: JS10000

Raw Sample Relinquished by: JS10000