



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127052

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/29/2023	11:35
2	CAL2	1	Water	NA	NA	20.7	7.00	08/29/2023	11:36
3	CAL3	1	Water	NA	NA	20.7	10.03	08/29/2023	11:37
4	ICV	1	Water	NA	NA	20.6	7.03	08/29/2023	11:40
5	CCV1	1	Water	NA	NA	20.7	2.02	08/29/2023	12:05
6	O4215-01	1	Water	NA	NA	20.2	4.32	08/29/2023	12:10
7	O4215-01DUP	1	Water	NA	NA	20.3	4.33	08/29/2023	12:11
8	CCV2	1	Water	NA	NA	20.4	12.01	08/29/2023	12:15

13127052

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity 04215

WorkList ID : 173126

Department : Wet-Chemistry

Date : 08-29-2023 11:26:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
04215-01	B AVE-DRUM-04-20230828	Water	Corrosivity	Cool 4 deg C	TETR16	L21	08/28/2023	9040C

Date/Time08/29/2311:30

Raw Sample Received by:jdcc

Raw Sample Relinquished by:afsm

Date/Time08/29/2314:00

Raw Sample Received by:afsm

Raw Sample Relinquished by:jdcc