



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/12/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:15
In Date: 08/29/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 08/30/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127056

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4186-01	EYF40	1	1.17	8.80	9.97	7.99	77.5	
O4186-02	EYF41	2	1.19	8.57	9.76	8.18	81.6	
O4186-03	EYF42	3	1.14	8.82	9.96	8.31	81.3	
O4186-04	EYF43	4	1.19	8.73	9.92	8.08	78.9	
O4186-05	EYF44	5	1.15	8.44	9.59	7.8	78.8	
O4186-07	E00D5	6	1.13	8.67	9.8	8.17	81.2	
O4186-08	E00D6	7	1.19	8.33	9.52	7.92	80.8	
O4186-09	EYEZ7	8	1.11	8.77	9.88	8.47	83.9	
O4186-10	EYEZ8	9	1.15	8.41	9.56	8.32	85.3	
O4186-11	EYEZ8D	10	1.15	8.41	9.56	8.32	85.3	
O4186-12	EYEZ8S	11	1.15	8.41	9.56	8.32	85.3	
O4186-13	EYEZ9	12	1.17	8.60	9.77	8.25	82.3	
O4186-14	EYF00	13	1.19	8.54	9.73	8.38	84.2	
O4186-15	EYF01	14	1.19	8.65	9.84	8.49	84.4	
O4186-16	EYF02	15	1.13	8.56	9.69	8.28	83.5	
O4186-17	EYF03	16	1.15	8.76	9.91	8.08	79.1	
O4186-18	EYF04	17	1.19	8.56	9.75	7.76	76.8	
O4186-19	EYF05	18	1.16	8.80	9.96	8.63	84.9	
O4186-20	EYF06	19	1.15	8.50	9.65	8.54	86.9	
O4186-21	EYF07	20	1.12	8.76	9.88	8.46	83.8	
O4186-22	EYF08	21	1.15	8.45	9.6	8.16	83.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$