


L8127060

SFAM01.1

INSTRUMENT ID: C1

Linear



A= 0.0000e+000
 B= 1.9652e-004 slope
 C= -4.5341e-003 y-intercept
 Rho= 0.9995431
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/1
0.00	0.000	0.026	0.026	154	0.000	154					=
0.05	0.050					501	0				-17
0.20	0.200	0.167	-0.033	871	0.0 %	871					-5
2.50	2.500	2.374	-0.126	12102	0.0 %	12102					-4 5
5.00	5.000	5.225	0.225	26612	0.0 %	26612					0 08
7.50	7.500	7.486	-0.014	38114	0.0 %	38114					-1 813123
10.0	10.000	9.923	-0.077	50515	0.0 %	50515					

LB127060
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	154	-		0 SFAM01.1	PPB	8/29/2023 16:06	Std
	0.2 S01	871	-		0.2 SFAM01.1	PPB	8/29/2023 16:08	Std
	2.5 S02	12102	-		2.5 SFAM01.1	PPB	8/29/2023 16:10	Std
	5 S03	26612	-		5 SFAM01.1	PPB	8/29/2023 16:12	Std
	7.5 S04	38114	-		7.5 SFAM01.1	PPB	8/29/2023 16:15	Std
	10 S05	50515	-		10 SFAM01.1	PPB	8/29/2023 16:17	Std
ICV038	ICV038	21374	4.1959 -		SFAM01.1	PPB	8/29/2023 16:20	SMPL
ICB038	ICB038	100	0.0151 -		SFAM01.1	PPB	8/29/2023 16:22	SMPL
CCV018	CCV018	26402	5.1841 -		SFAM01.1	PPB	8/29/2023 16:24	SMPL
CCB018	CCB018	109	0.0169 -		SFAM01.1	PPB	8/29/2023 16:27	SMPL
PB155188BL	PBW188	98	0.0147 -		SFAM01.1	PPB	8/29/2023 16:29	SMPL
O4041-01	MEZYN5	51	0.0055 -		SFAM01.1	PPB	8/29/2023 16:31	SMPL
O4041-02	MEZYX2	64	0.008 -		SFAM01.1	PPB	8/29/2023 16:33	SMPL
O4041-03	MEZYN6	75	0.0102 -		SFAM01.1	PPB	8/29/2023 16:36	SMPL
O4041-04	MEZYN7	76	0.0104 -		SFAM01.1	PPB	8/29/2023 16:38	SMPL
PB155193BL	PBW193	129	0.0208 -		SFAM01.1	PPB	8/29/2023 16:40	SMPL
O4061-01	MC0AG1	266	0.0477 -		SFAM01.1	PPB	8/29/2023 16:42	SMPL
O4061-02	MC0AG5	306	0.0556 -		SFAM01.1	PPB	8/29/2023 16:45	SMPL
O4061-03	MC0AG6	250	0.0446 -		SFAM01.1	PPB	8/29/2023 16:47	SMPL
O4061-04	MC0AG6D	228	0.0403 -		SFAM01.1	PPB	8/29/2023 16:49	SMPL
O4061-05	MC0AG6S	6357	1.2448 -		SFAM01.1	PPB	8/29/2023 16:51	SMPL
PB155194BL	PBW194	-15	-0.0075 -		SFAM01.1	PPB	8/29/2023 16:56	SMPL
O4062-01	MC0AG2	593	0.112 -		SFAM01.1	PPB	8/29/2023 16:59	SMPL
O4062-02	MC0AG3	375	0.0692 -		SFAM01.1	PPB	8/29/2023 17:01	SMPL
O4062-03	MC0AG4	251	0.0448 -		SFAM01.1	PPB	8/29/2023 17:04	SMPL
O4062-04	MC0AG4D	127	0.0204 -		SFAM01.1	PPB	8/29/2023 17:06	SMPL
O4062-05	MC0AG4S	4959	0.97 -		SFAM01.1	PPB	8/29/2023 17:08	SMPL
CCV019	CCV019	26359	5.1756 -		SFAM01.1	PPB	8/29/2023 17:10	SMPL
CCB019	CCB019	131	0.0212 -		SFAM01.1	PPB	8/29/2023 17:13	SMPL