



PERCENT SOLID

Supervisor: Sohil  
Analyst: JIGNESH  
Date: 8/30/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:15  
In Date: 08/29/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:14  
Out Date: 08/30/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB127062

| Lab ID   | Client SampleID      | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|----------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| O4210-01 | TR-04-082923         | 1      | 1.19           | 8.46         | 9.65                    | 8.84                      | 90.4    |                 |
| O4210-02 | TR-04-082923-E2      | 2      | 1.19           | 8.43         | 9.62                    | 8.51                      | 86.8    |                 |
| O4212-01 | TOPSOIL-OU3-DENALI-2 | 3      | 1.19           | 8.64         | 9.83                    | 6.75                      | 64.4    |                 |
| O4212-03 | TOPSOIL-OU3-DENALI-3 | 4      | 1.13           | 8.45         | 9.58                    | 6.32                      | 61.4    |                 |
| O4217-01 | BU-04-082923         | 5      | 1.15           | 8.53         | 9.68                    | 9.43                      | 97.1    |                 |
| O4218-01 | COMP-1               | 9      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CONCRETE SAMPLE |
| O4218-02 | COMP-2               | 10     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CONCRETE SAMPLE |
| O4219-01 | 72-11939             | 11     | 1.12           | 8.72         | 9.84                    | 8.27                      | 82.0    |                 |
| O4219-02 | 72-11939             | 12     | 1.13           | 8.63         | 9.76                    | 8.62                      | 86.8    |                 |
| O4219-03 | ETGI-350             | 13     | 1.11           | 8.73         | 9.84                    | 8.4                       | 83.5    |                 |
| O4219-04 | ETGI-350             | 14     | 1.12           | 8.63         | 9.75                    | 8.41                      | 84.5    |                 |
| O4221-01 | AU-5-082923          | 6      | 1.19           | 8.54         | 9.73                    | 9.33                      | 95.3    |                 |
| O4224-01 | EO-03-8292023        | 7      | 1.18           | 8.62         | 9.8                     | 8.91                      | 89.7    |                 |
| O4224-02 | EO-03-8292023        | 8      | 1.18           | 8.38         | 9.56                    | 8.77                      | 90.6    |                 |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$