



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB127063

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/29/2023	17:05
2	CAL2	1	Water	NA	NA	20.7	7.00	08/29/2023	17:06
3	CAL3	1	Water	NA	NA	20.6	10.02	08/29/2023	17:10
4	ICV	1	Water	NA	NA	20.7	7.01	08/29/2023	17:15
5	CCV1	1	Water	NA	NA	20.7	2.02	08/29/2023	17:20
6	O4222-01	1	Water	NA	NA	23.4	5.69	08/29/2023	17:25
7	O4222-01DUP	1	Water	NA	NA	23.5	5.70	08/29/2023	17:26
8	O4222-02	1	Water	NA	NA	23.5	7.40	08/29/2023	17:30
9	CCV2	1	Water	NA	NA	20.6	12.02	08/29/2023	17:32

WORKLIST(Hardcopy Internal Chain)

18121063

WorkList Name : ph 04217

WorkList ID : 173142

Department : Wet-Chemistry

Date : 08-29-2023 16:45:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4222-01	A 1039	Water	pH	Cool 4 deg C	PSEG03	N21	08/29/2023	9040C
O4222-02	1040	Water	pH	Cool 4 deg C	PSEG03	N21	08/29/2023	9040C

Date/Time 08.29.23 17:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC

Date/Time 08.29.23 18:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC