

7470A


LB127073

INSTRUMENTAL ID: CV1

Linear

 μ Abs.:

58009



A= 0.0000e+000

B= 1.7243e-004

C= -7.0852e-002

Rho= 0.9993146

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Good
0.0	0.000	-0.045	-0.045	149	0.000	149	0				1
0.2	0.200	0.179	-0.021	1450	0.0 %	1450					1
2.5	2.500	2.436	-0.064	14539	0.0 %	14539					3
5.0	5.000	5.300	0.300	31151	0.0 %	31151					6
7.5	7.500	7.398	-0.102	43315	0.0 %	43315					1
10.0	10.000	9.932	-0.068	58009	0.0 %	58009					1

LB127073 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Aug 2023 12:05:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	149	-	0.0000	7470A	PPB	30 Aug 2023 12:15:42	S	Std
0.2 - 1	50.2	1450	-	0.2000	7470A	PPB	30 Aug 2023 12:17:58	S	Std
2.5 - 1	52.5	14539	-	2.5000	7470A	PPB	30 Aug 2023 12:20:15	S	Std
5.0 - 1	55	31151	-	5.0000	7470A	PPB	30 Aug 2023 12:22:32	S	Std
7.5 - 1	58.5	43315	-	7.5000	7470A	PPB	30 Aug 2023 12:24:48	S	Std
10.0 - 1	510	58009	-	10.0000	7470A	PPB	30 Aug 2023 12:27:04	S	Std
ICV39 - 1	ICV39	24085	4.0821	-	7470A	PPB	30 Aug 2023 12:31:42	U	SMPL
ICB39 - 1	ICB39	55	-0.0614	-	7470A	PPB	30 Aug 2023 12:33:57	U	SMPL
CCV20 - 1	CCV20	30380	5.1675	-	7470A	PPB	30 Aug 2023 12:36:12	U	SMPL
CCB20 - 1	CCB20	61	-0.0603	-	7470A	PPB	30 Aug 2023 12:38:27	U	SMPL
CRA - 1	CRA	1418	0.1737	-	7470A	PPB	30 Aug 2023 12:40:41	U	SMPL
HighStd - 1	HighStd	54863	9.3891	-	7470A	PPB	30 Aug 2023 12:42:56	U	SMPL
ChkStd - 1	ChkStd	39868	6.8035	-	7470A	PPB	30 Aug 2023 12:45:11	U	SMPL
PB155186BL - 1	PBW	86	-0.0560	-	7470A	PPB	30 Aug 2023 12:47:26	U	SMPL
PB155186BS - 1	LCSW	21929	3.7103	-	7470A	PPB	30 Aug 2023 12:49:42	U	SMPL
O4196-02 - 1	OR-34-COMP-01	127	-0.0490	-	7470A	PPB	30 Aug 2023 12:51:58	U	SMPL
O4196-02DUP - 1	OR-34-COMP-01DUP	144	-0.0460	-	7470A	PPB	30 Aug 2023 12:54:15	U	SMPL
O4196-02MS - 1	OR-34-COMP-01MS	21744	3.6784	-	7470A	PPB	30 Aug 2023 12:56:32	U	SMPL
O4196-02MSD - 1	OR-34-COMP-01MSD	22589	3.8241	-	7470A	PPB	30 Aug 2023 12:58:49	U	SMPL
O4196-09 - 1	OR-345-COMP-02	64	-0.0598	-	7470A	PPB	30 Aug 2023 13:01:04	U	SMPL
CCV21 - 1	CCV21	31096	5.2910	-	7470A	PPB	30 Aug 2023 13:03:19	U	SMPL
CCB21 - 1	CCB21	67	-0.0593	-	7470A	PPB	30 Aug 2023 13:05:34	U	SMPL
O4196-16 - 1	OR-345-COMP-03	120	-0.0502	-	7470A	PPB	30 Aug 2023 13:07:49	U	SMPL
O4196-23 - 1	OR-345-COMP-04	126	-0.0491	-	7470A	PPB	30 Aug 2023 13:10:04	U	SMPL
O4196-30 - 1	OR-345-COMP-05	98	-0.0540	-	7470A	PPB	30 Aug 2023 13:12:19	U	SMPL
O4197-02 - 1	OR-309-OK-370-COMP-20	131	-0.0483	-	7470A	PPB	30 Aug 2023 13:14:35	U	SMPL
O4197-09 - 1	OR-309-OK-370-COMP-21	99	-0.0538	-	7470A	PPB	30 Aug 2023 13:16:51	U	SMPL
O4197-16 - 1	OR-309-OK-370-COMP-22	111	-0.0517	-	7470A	PPB	30 Aug 2023 13:19:06	U	SMPL
O4197-23 - 1	OR-309-OK-370-COMP-23	99	-0.0538	-	7470A	PPB	30 Aug 2023 13:21:23	U	SMPL
O4198-02 - 1	OR-309-OK-370-COMP-24	105	-0.0527	-	7470A	PPB	30 Aug 2023 13:23:40	U	SMPL
O4198-08 - 1	OR-309-OK-370-COMP-25	83	-0.0565	-	7470A	PPB	30 Aug 2023 13:25:56	U	SMPL
O4198-15 - 1	OR-309-OK-370-COMP-26	139	-0.0469	-	7470A	PPB	30 Aug 2023 13:28:12	U	SMPL
CCV22 - 1	CCV22	31072	5.2868	-	7470A	PPB	30 Aug 2023 13:30:27	U	SMPL
CCB22 - 1	CCB22	113	-0.0514	-	7470A	PPB	30 Aug 2023 13:32:43	U	SMPL
O4200-01 - 1	COMP-1	196	-0.0371	-	7470A	PPB	30 Aug 2023 13:34:58	U	SMPL
O4200-02 - 1	COMP-2	161	-0.0431	-	7470A	PPB	30 Aug 2023 13:37:14	U	SMPL
O4200-03 - 1	COMP-3	182	-0.0395	-	7470A	PPB	30 Aug 2023 13:39:29	U	SMPL
O4200-04 - 1	8372	124	-0.0495	-	7470A	PPB	30 Aug 2023 13:41:45	U	SMPL
PB155152TB - 1	PB155152TB	104	-0.0529	-	7470A	PPB	30 Aug 2023 13:44:00	U	SMPL
O4196-02LX5 - 1		64	-0.0598	-	7470A	PPB	30 Aug 2023 13:46:16	U	SMPL
O4196-02A - 1		24991	4.2383	-	7470A	PPB	30 Aug 2023 13:48:31	U	SMPL
CCV23 - 1	CCV23	28354	4.8182	-	7470A	PPB	30 Aug 2023 13:50:47	U	SMPL
CCB23 - 1	CCB23	74	-0.0581	-	7470A	PPB	30 Aug 2023 13:53:02	U	SMPL
							30 Aug 2023 14:00:11	U	SMPL
							30 Aug 2023 14:02:28	U	SMPL