



PERCENT SOLID

Supervisor: Sohil

Analyst: JIGNESH

Date: 8/31/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 14:35

In Date: 08/30/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:47

Out Date: 08/31/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB127077

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4191-01	EYET3	1	1.12	8.77	9.89	8.2	80.7	
O4191-02	EYET4	2	1.16	8.44	9.6	7.88	79.6	
O4191-03	EYET5	3	1.19	8.61	9.8	8.01	79.2	
O4191-04	EYET6	4	1.15	8.62	9.77	7.83	77.5	
O4191-05	EYET7	5	1.19	8.79	9.98	8.14	79.1	
O4191-06	EYET8	6	1.19	8.43	9.62	7.79	78.3	
O4191-07	EYET9	7	1.14	8.80	9.94	8.41	82.6	
O4191-08	EYEW0	8	1.15	8.40	9.55	7.79	79.0	
O4191-09	EYEW1	9	1.1	8.79	9.89	8.56	84.9	
O4191-10	EYEW2	10	1.15	8.38	9.53	8.16	83.7	
O4191-11	EYEW3	11	1.16	8.48	9.64	8.05	81.3	
O4191-12	EYEW4	12	1.16	8.50	9.66	7.99	80.4	
O4191-13	EYEW5	13	1.19	8.51	9.7	8.08	81.0	
O4191-14	EYEW6	14	1.11	8.68	9.79	8.23	82.0	
O4191-15	EYEW7	15	1.16	8.48	9.64	7.99	80.5	
O4191-16	EYEW8	16	1.19	8.34	9.53	8.08	82.6	
O4191-17	EYEW9	17	1.17	8.81	9.98	8.42	82.3	
O4191-18	EYEX0	18	1.17	8.58	9.75	8.32	83.3	
O4191-19	EYEX1	19	1.15	8.67	9.82	8.24	81.8	
O4191-20	EYEX1D	20	1.15	8.67	9.82	8.24	81.8	
O4191-21	EYEX1S	21	1.15	8.67	9.82	8.24	81.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$