


LB127079

SFAM01.1

INSTRUMENT ID: CVL



Linear

A= 0.0000e+000

B= 1.7561e-004 slope

C= -8.6596e-002 y-intercept

Rho= 0.9999592

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90.0
0.00	0.000	0.002	0.002	507	0.000	507					
0.05	0.050					501	0				
0.20	0.200	0.185	-0.015	1546	0.0 %	1546					
2.50	2.500	2.499	-0.001	14723	0.0 %	14723					
5.00	5.000	4.995	-0.005	28938	0.0 %	28938					
7.50	7.500	7.565	0.065	43573	0.0 %	43573					
10.0	10.000	9.953	-0.047	57173	0.0 %	57173					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	507	-		0 SFAM01.1	PPB	8/30/2023 14:31	Std
	0.2 S01	1546	-		0.2 SFAM01.1	PPB	8/30/2023 14:33	Std
	2.5 S02	14723	-		2.5 SFAM01.1	PPB	8/30/2023 14:35	Std
	5 S03	28938	-		5 SFAM01.1	PPB	8/30/2023 14:38	Std
	7.5 S04	43573	-		7.5 SFAM01.1	PPB	8/30/2023 14:40	Std
	10 S05	57173	-		10 SFAM01.1	PPB	8/30/2023 14:42	Std
ICV040	ICV040	22703	3.9002 -		SFAM01.1	PPB	8/30/2023 14:48	SMPL
ICB040	ICB040	472	-0.0037 -		SFAM01.1	PPB	8/30/2023 14:50	SMPL
CCV024	CCV024	28940	4.9955 -		SFAM01.1	PPB	8/30/2023 14:52	SMPL
CCB024	CCB024	474	-0.0034 -		SFAM01.1	PPB	8/30/2023 14:54	SMPL
PB155227BL	PBS227	55	-0.0769 -		SFAM01.1	PPB	8/30/2023 14:57	SMPL
O4083-01	GCED3	1085	0.1039 -		SFAM01.1	PPB	8/30/2023 14:59	SMPL
O4083-02	GCED3D	964	0.0827 -		SFAM01.1	PPB	8/30/2023 15:01	SMPL
O4083-03	GCED3S	16163	2.7518 -		SFAM01.1	PPB	8/30/2023 15:03	SMPL
O4083-04	GCED4	1275	0.1373 -		SFAM01.1	PPB	8/30/2023 15:06	SMPL
O4083-05	GCED5	1464	0.1705 -		SFAM01.1	PPB	8/30/2023 15:08	SMPL
O4083-06	GCED6	661	0.0295 -		SFAM01.1	PPB	8/30/2023 15:10	SMPL
O4083-07	GCED7	1299	0.1415 -		SFAM01.1	PPB	8/30/2023 15:12	SMPL
O4083-08	GCED8	816	0.0567 -		SFAM01.1	PPB	8/30/2023 15:15	SMPL
O4083-09	GCED9	1064	0.1003 -		SFAM01.1	PPB	8/30/2023 15:17	SMPL
O4083-10	GCEE0	816	0.0567 -		SFAM01.1	PPB	8/30/2023 15:19	SMPL
O4083-11	GCEE1	728	0.0412 -		SFAM01.1	PPB	8/30/2023 15:21	SMPL
O4083-12	GCEE2	1466	0.1708 -		SFAM01.1	PPB	8/30/2023 15:24	SMPL
O4083-13	GCEE3	650	0.0275 -		SFAM01.1	PPB	8/30/2023 15:26	SMPL
O4083-14	GCEE4	932	0.0771 -		SFAM01.1	PPB	8/30/2023 15:28	SMPL
O4083-15	GCEE5	564	0.0124 -		SFAM01.1	PPB	8/30/2023 15:30	SMPL
O4083-16	GCEE6	800	0.0539 -		SFAM01.1	PPB	8/30/2023 15:33	SMPL
O4083-17	GCEE7	1081	0.1032 -		SFAM01.1	PPB	8/30/2023 15:35	SMPL
O4083-18	GCEE8	760	0.0469 -		SFAM01.1	PPB	8/30/2023 15:37	SMPL
O4083-19	GCEE9	707	0.0376 -		SFAM01.1	PPB	8/30/2023 15:40	SMPL
CCV025	CCV025	27729	4.7828 -		SFAM01.1	PPB	8/30/2023 15:42	SMPL
CCB025	CCB025	426	-0.0118 -		SFAM01.1	PPB	8/30/2023 15:44	SMPL
O4083-20	GCEF0	1246	0.1322 -		SFAM01.1	PPB	8/30/2023 15:46	SMPL
O4083-21	GCEF1	742	0.0437 -		SFAM01.1	PPB	8/30/2023 15:49	SMPL
PB155228BL	PBS228	97	-0.0696 -		SFAM01.1	PPB	8/30/2023 15:51	SMPL
O4084-01	GCEF2	1202	0.1245 -		SFAM01.1	PPB	8/30/2023 15:53	SMPL
O4084-02	GCEF3	950	0.0802 -		SFAM01.1	PPB	8/30/2023 15:55	SMPL
O4084-03	GCEF3D	973	0.0843 -		SFAM01.1	PPB	8/30/2023 15:58	SMPL
O4084-04	GCEF3S	15633	2.6587 -		SFAM01.1	PPB	8/30/2023 16:00	SMPL
O4084-05	GCEF4	4968	0.7858 -		SFAM01.1	PPB	8/30/2023 16:02	SMPL
O4084-06	GCEF5	750	0.0451 -		SFAM01.1	PPB	8/30/2023 16:04	SMPL
O4084-07	GCEF6	964	0.0827 -		SFAM01.1	PPB	8/30/2023 16:07	SMPL
O4084-08	GCEF7	834	0.0599 -		SFAM01.1	PPB	8/30/2023 16:09	SMPL
O4084-09	GCEF8	1045	0.0969 -		SFAM01.1	PPB	8/30/2023 16:11	SMPL
O4084-10	GCEF9	1190	0.1224 -		SFAM01.1	PPB	8/30/2023 16:13	SMPL
O4084-11	GCEG0	2374	0.3303 -		SFAM01.1	PPB	8/30/2023 16:16	SMPL

LB127079
INSTRUMENT ID : CV1

O4084-12	GCEG1	4174	0.6464 -	SFAM01.1	PPB	8/30/2023 16:18 SMPL
O4084-13	GCEG2	712	0.0384 -	SFAM01.1	PPB	8/30/2023 16:20 SMPL
O4084-14	GCEG3	601	0.0189 -	SFAM01.1	PPB	8/30/2023 16:22 SMPL
O4084-15	GCEG4	869	0.066 -	SFAM01.1	PPB	8/30/2023 16:25 SMPL
O4084-16	GCEG5	972	0.0841 -	SFAM01.1	PPB	8/30/2023 16:27 SMPL
O4084-17	GCEG6	953	0.0808 -	SFAM01.1	PPB	8/30/2023 16:29 SMPL
CCV026	CCV026	27479	4.7389 -	SFAM01.1	PPB	8/30/2023 16:32 SMPL
CCB026	CCB026	431	-0.0109 -	SFAM01.1	PPB	8/30/2023 16:34 SMPL
O4084-18	GCEG7	896	0.0707 -	SFAM01.1	PPB	8/30/2023 16:36 SMPL
O4084-19	GCEG8	1979	0.2609 -	SFAM01.1	PPB	8/30/2023 16:38 SMPL
O4084-20	GCEG9	921	0.0751 -	SFAM01.1	PPB	8/30/2023 16:41 SMPL
O4084-21	GCEH0	942	0.0788 -	SFAM01.1	PPB	8/30/2023 16:43 SMPL
PB155229BL	PBS229	53	-0.0773 -	SFAM01.1	PPB	8/30/2023 16:45 SMPL
O4085-01	GCEH1	774	0.0493 -	SFAM01.1	PPB	8/30/2023 16:47 SMPL
O4085-02	GCEH2	1258	0.1343 -	SFAM01.1	PPB	8/30/2023 16:50 SMPL
O4085-03	GCEH3	1342	0.1491 -	SFAM01.1	PPB	8/30/2023 16:52 SMPL
O4085-04	GCEH3D	1253	0.1334 -	SFAM01.1	PPB	8/30/2023 16:54 SMPL
O4085-05	GCEH3S	15149	2.5737 -	SFAM01.1	PPB	8/30/2023 16:56 SMPL
O4085-06	GCEH4	1110	0.1083 -	SFAM01.1	PPB	8/30/2023 16:59 SMPL
O4085-07	GCEH5	1840	0.2365 -	SFAM01.1	PPB	8/30/2023 17:01 SMPL
O4085-08	GCEH6	3095	0.4569 -	SFAM01.1	PPB	8/30/2023 17:03 SMPL
O4085-09	GCEH7	1468	0.1712 -	SFAM01.1	PPB	8/30/2023 17:06 SMPL
O4085-10	GCEH8	857	0.0639 -	SFAM01.1	PPB	8/30/2023 17:08 SMPL
O4085-11	GCEH9	1387	0.157 -	SFAM01.1	PPB	8/30/2023 17:10 SMPL
O4085-12	GCEJ0	1548	0.1852 -	SFAM01.1	PPB	8/30/2023 17:12 SMPL
O4085-13	GCEJ1	1079	0.1029 -	SFAM01.1	PPB	8/30/2023 17:15 SMPL
O4085-14	GCEJ2	1390	0.1575 -	SFAM01.1	PPB	8/30/2023 17:17 SMPL
O4085-15	GCEJ3	1204	0.1248 -	SFAM01.1	PPB	8/30/2023 17:19 SMPL
CCV027	CCV027	27478	4.7388 -	SFAM01.1	PPB	8/30/2023 17:21 SMPL
CCB027	CCB027	426	-0.0118 -	SFAM01.1	PPB	8/30/2023 17:24 SMPL
O4085-16	GCEJ4	1012	0.0911 -	SFAM01.1	PPB	8/30/2023 17:26 SMPL
O4085-17	GCEJ5	1215	0.1268 -	SFAM01.1	PPB	8/30/2023 17:28 SMPL
O4085-18	GCEJ6	29009	5.0076 -	SFAM01.1	PPB	8/30/2023 17:33 SMPL
CCV028	CCV028	27215	4.6926 -	SFAM01.1	PPB	8/30/2023 17:40 SMPL
CCB028	CCB028	420	-0.0128 -	SFAM01.1	PPB	8/30/2023 17:42 SMPL