



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 9/1/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 13:25

In Date: 08/31/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:44

Out Date: 09/01/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB127088

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4176-01	MEZYK2	1	1.14	8.66	9.8	8.94	90.1	
O4176-02	MEZYK3	2	1.16	8.49	9.65	8.71	88.9	
O4176-03	MEZYK4	3	1.12	8.57	9.69	8.42	85.2	
O4176-05	MEZYC9	4	1.19	8.65	9.84	8.00	78.7	
O4176-06	MEZYC9D	5	1.19	8.65	9.84	8.00	78.7	
O4176-07	MEZYC9S	6	1.19	8.65	9.84	8.00	78.7	
O4176-08	MEZYD0	7	1.16	8.67	9.83	8.79	88.0	
O4176-09	MEZYD1	8	1.18	8.73	9.91	9.09	90.6	
O4176-10	MEZYD2	9	1.15	8.82	9.97	9.18	91.0	
O4176-11	MEZYD3	10	1.17	8.58	9.75	8.7	87.8	
O4176-12	MEZYD4	11	1.19	8.41	9.6	7.96	80.5	
O4176-13	MEZYL5	12	1.18	8.56	9.74	8.67	87.5	
O4176-15	MEZYD8	13	1.11	8.74	9.85	8.66	86.4	
O4176-16	MEZYD9	14	1.16	8.83	9.99	9.04	89.2	
O4176-17	MEZYE0	15	1.12	8.70	9.82	8.55	85.4	
O4176-18	MEZYJ9	16	1.18	8.69	9.87	9.06	90.7	
O4176-19	MEZYK0	17	1.15	8.81	9.96	9.42	93.9	
O4176-20	MEZYK1	18	1.19	8.44	9.63	6.45	62.3	
O4176-21	MEZYL8	19	1.19	8.44	9.63	8.8	90.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$