



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 9/1/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 13:25

In Date: 08/31/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:44

Out Date: 09/01/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: % SOLIDS-OVEN

QC:LB127089

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4175-01	EZYG5	1	1.15	8.38	9.53	8.77	90.9	
O4175-02	EZYG6	2	1.18	8.75	9.93	9.03	89.7	
O4175-03	EZYG7	3	1.19	8.73	9.92	8.49	83.6	
O4175-05	EZYGX0	4	1.17	8.55	9.72	8.8	89.2	
O4175-06	EZYGX3	5	1.19	8.35	9.54	8.53	87.9	
O4175-07	EZYGX4	6	1.11	8.56	9.67	7.95	79.9	
O4175-08	EZYK2	7	1.14	8.66	9.8	8.94	90.1	
O4175-09	EZYK3	8	1.16	8.49	9.65	8.71	88.9	
O4175-10	EZYK4	9	1.12	8.57	9.69	8.42	85.2	
O4175-13	VHBLK002	10	1.00	1.00	2.00	2.00	100.0	vhblk
O4175-14	EZYC9	11	1.19	8.65	9.84	8.00	78.7	
O4175-15	EZYC9MS	12	1.19	8.65	9.84	8.00	78.7	
O4175-16	EZYC9MSD	13	1.19	8.65	9.84	8.00	78.7	
O4175-17	EZYD0	14	1.16	8.67	9.83	8.79	88.0	
O4175-18	EZYD1	15	1.18	8.73	9.91	9.09	90.6	
O4175-19	EZYD2	16	1.15	8.82	9.97	9.18	91.0	
O4175-20	EZYD3	17	1.17	8.58	9.75	8.7	87.8	
O4175-21	EZYD4	18	1.19	8.41	9.6	7.96	80.5	
O4175-22	EZYW5	19	1.18	8.56	9.74	8.67	87.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$