

LB127090


7471B

INSTRUMENT ID: RV1

Linear

 μ Abs.:

65023



A= 0.0000e+000

B= 1.5468e-004

C= -1.9958e-001

Rho= 0.9995815

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Handwritten
0.0	0.000	-0.081	-0.081	764	0.000	764					1.5
0.2	0.200	0.225	0.025	2746	0.0 %	2746					13
2.5	2.500	2.550	0.050	17773	0.0 %	17773					2
5.0	5.000	4.957	-0.043	33338	0.0 %	33338					1
7.5	7.500	7.692	0.192	51016	0.0 %	51016					3
10.0	10.000	9.858	-0.142	65023	0.0 %	65023					1

LB127090

Method: 7471B

Operator: Admin

Date of Analysis: 31 Aug 2023 09:47:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	764	-	0.0000	7471B	PPB	31 Aug 2023 09:56:01	S	Std
0.2 - 1	50.2	2746	-	0.2000	7471B	PPB	31 Aug 2023 09:58:18	S	Std
2.5 - 1	52.5	17773	-	2.5000	7471B	PPB	31 Aug 2023 10:00:35	S	Std
5.0 - 1	55	33338	-	5.0000	7471B	PPB	31 Aug 2023 10:02:51	S	Std
7.5 - 1	57.5	51016	-	7.5000	7471B	PPB	31 Aug 2023 10:05:08	S	Std
10.0 - 1	5/0	65023	-	10.0000	7471B	PPB	31 Aug 2023 10:09:59	S	Std
ICV41 - 1	ICV41	27985	4.1291	-	7471B	PPB	31 Aug 2023 10:12:53	U	SMPL
ICB41 - 1	ICB41	781	-0.0788	-	7471B	PPB	31 Aug 2023 10:17:47	U	SMPL
CCV29 - 1	CCV29	33072	4.9160	-	7471B	PPB	31 Aug 2023 10:20:01	U	SMPL
CCB29 - 1	CCB29	745	-0.0843	-	7471B	PPB	31 Aug 2023 10:22:16	U	SMPL
CRA - 1	CRA	2701	0.2182	-	7471B	PPB	31 Aug 2023 10:24:33	U	SMPL
HighStd - 1	HighStd	64847	9.8309	-	7471B	PPB	31 Aug 2023 10:26:48	U	SMPL
ChkStd - 1	ChkStd	47813	7.1961	-	7471B	PPB	31 Aug 2023 10:29:03	U	SMPL
PB155244BL - 1	PBS	919	-0.0574	-	7471B	PPB	31 Aug 2023 10:33:58	U	SMPL
PB155244BS - 1	LCSS	27439	4.0447	-	7471B	PPB	31 Aug 2023 10:36:13	U	SMPL
O4210-01 - 1	TR-04-082923	4678	0.5240	-	7471B	PPB	31 Aug 2023 10:38:29	U	SMPL
O4210-01DUP - 1	TR-04-082923DUP	4913	0.5604	-	7471B	PPB	31 Aug 2023 10:40:45	U	SMPL
O4210-01MS - 1	TR-04-082923MS	33441	4.9730	-	7471B	PPB	31 Aug 2023 10:43:02	U	SMPL
O4210-01MSD - 1	TR-04-082923MSD	31569	4.6835	-	7471B	PPB	31 Aug 2023 10:47:53	U	SMPL
O4212-01 - 1	TOPSOIL-OU3-DENALI-2	2926	0.2530	-	7471B	PPB	31 Aug 2023 10:50:08	U	SMPL
CCV30 - 1	CCV30	32291	4.7952	-	7471B	PPB	31 Aug 2023 10:52:24	U	SMPL
CCB30 - 1	CCB30	738	-0.0854	-	7471B	PPB	31 Aug 2023 10:54:39	U	SMPL
O4212-03 - 1	TOPSOIL-OU3-DENALI-3	4870	0.5537	-	7471B	PPB	31 Aug 2023 10:56:55	U	SMPL
O4217-01 - 1	BU-04-082923	5605	0.6674	-	7471B	PPB	31 Aug 2023 10:59:10	U	SMPL
O4219-02 - 1	72-11939	13373	1.8689	-	7471B	PPB	31 Aug 2023 11:01:25	U	SMPL
O4219-04 - 1	ETGI-350	6701	0.8369	-	7471B	PPB	31 Aug 2023 11:03:41	U	SMPL
O4221-01 - 1	AU-5-082923	4201	0.4502	-	7471B	PPB	31 Aug 2023 11:05:57	U	SMPL
O4224-01 - 1	EO-03-8292023	3633	0.3624	-	7471B	PPB	31 Aug 2023 11:08:12	U	SMPL
O4244-02 - 1	BIN-BLOCK	1116	-0.0270	-	7471B	PPB	31 Aug 2023 11:10:29	U	SMPL
O4210-01LX5 - 1		1004	-0.0443	-	7471B	PPB	31 Aug 2023 11:12:46	U	SMPL
O4210-01A - 1		29800	4.4098	-	7471B	PPB	31 Aug 2023 11:15:02	U	SMPL
CCV31 - 1	CCV31	32549	4.8351	-	7471B	PPB	31 Aug 2023 11:17:18	U	SMPL
CCB31 - 1	CCB31	805	-0.0751	-	7471B	PPB	31 Aug 2023 11:22:11	U	SMPL