

LB127094


7470A

INSTRUMENT-ID: CV1

Linear

 μ Abs.:

5278E



A= 0.0000e+000

B= 1.8825e-004

C= -2.0561e-001

Rho= 0.9987497

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.0	0.000	-0.058	-0.058	784	0.000	784	0				1
0.2	0.200	0.192	-0.008	2111	0.0 %	2111					-4
2.5	2.500	2.530	0.030	14532	0.0 %	14532					1
5.0	5.000	4.951	-0.049	27391	0.0 %	27391					-1
7.5	7.500	7.854	0.354	42812	0.0 %	42812					5
10.0	10.000	9.731	-0.269	52785	0.0 %	52785					3

LB127094 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Aug 2023 13:41:35

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	784	-	0.0000	7470A	PPB	31 Aug 2023 13:44:54	S	Std
0.2 - 1	50.2	2111	-	0.2000	7470A	PPB	31 Aug 2023 13:47:10	S	Std
2.5 - 1	52.5	14532	-	2.5000	7470A	PPB	31 Aug 2023 13:49:28	S	Std
5.0 - 1	55	27391	-	5.0000	7470A	PPB	31 Aug 2023 13:51:44	S	Std
7.5 - 1	57.5	42812	-	7.5000	7470A	PPB	31 Aug 2023 13:54:01	S	Std
10.0 - 1	57.5	52785	-	10.0000	7470A	PPB	31 Aug 2023 13:56:17	S	Std
ICV42 - 1	ICV42	22131	3.9606	-	7470A	PPB	31 Aug 2023 14:01:57	U	SMPL
ICB42 - 1	ICB42	789	-0.0571	-	7470A	PPB	31 Aug 2023 14:04:11	U	SMPL
CCV32 - 1	CCV32	27313	4.9362	-	7470A	PPB	31 Aug 2023 14:06:26	U	SMPL
CCB32 - 1	CCB32	796	-0.0558	-	7470A	PPB	31 Aug 2023 14:08:41	U	SMPL
CRA - 1	CRA	2027	0.1760	-	7470A	PPB	31 Aug 2023 14:10:55	U	SMPL
HighStd - 1	HighStd	52530	9.6834	-	7470A	PPB	31 Aug 2023 14:13:11	U	SMPL
ChkStd - 1	ChkStd	39861	7.2984	-	7470A	PPB	31 Aug 2023 14:15:26	U	SMPL
PB155233BL - 1	PBW	606	-0.0915	-	7470A	PPB	31 Aug 2023 14:17:42	U	SMPL
PB155233BS - 1	LCSW	20289	3.6139	-	7470A	PPB	31 Aug 2023 14:22:57	U	SMPL
O4212-02 - 1	TOPSOIL-OU3-DENALI-2	661	-0.0812	-	7470A	PPB	31 Aug 2023 14:25:13	U	SMPL
O4212-04 - 1	TOPSOIL-OU3-DENALI-3	661	-0.0812	-	7470A	PPB	31 Aug 2023 14:27:30	U	SMPL
O4212-04DUP - 1	TOPSOIL-OU3-DENALI-3DUP	673	-0.0789	-	7470A	PPB	31 Aug 2023 14:32:55	U	SMPL
O4212-04MS - 1	TOPSOIL-OU3-DENALI-3MS	20620	3.6762	-	7470A	PPB	31 Aug 2023 14:35:12	U	SMPL
O4212-04MSD - 1	TOPSOIL-OU3-DENALI-3MSD	21590	3.8588	-	7470A	PPB	31 Aug 2023 14:42:49	U	SMPL
CCV33 - 1	CCV33	27659	5.0013	-	7470A	PPB	31 Aug 2023 14:45:03	U	SMPL
CCB33 - 1	CCB33	826	-0.0501	-	7470A	PPB	31 Aug 2023 14:47:18	U	SMPL
PB155187TB - 1	PB155187TB	613	-0.0902	-	7470A	PPB	31 Aug 2023 14:49:33	U	SMPL
O4212-04LX5 - 1		156	-0.1762	-	7470A	PPB	31 Aug 2023 14:51:48	U	SMPL
O4212-04A - 1		25526	4.5998	-	7470A	PPB	31 Aug 2023 14:54:03	U	SMPL
CCV34 - 1	CCV34	28657	5.1892	-	7470A	PPB	31 Aug 2023 14:56:19	U	SMPL
CCB34 - 1	CCB34	736	-0.0671	-	7470A	PPB	31 Aug 2023 14:58:34	U	SMPL