



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: JIGNESH

Run Number: LB127102

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/31/2023	12:50
2	O4231-01		1	11	100.00	>212.0	>212.0	08/31/2023	13:20
3	O4231-01DUP		1	12	100.00	>212.0	>212.0	08/31/2023	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6/27/02

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP S-8-30 WorkList ID : 173192 Department : Wet-Chemistry Date : 08-30-2023 19:35:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4231-01	0706	Solid	Flash Point	Cool 4 deg C	PSEG03	N11	08/30/2023	1010B

Date/Time 08/31/2023 12:45
Raw Sample Received by: BM CWL
Raw Sample Relinquished by: JH (WCL)

Date/Time 08/31/2023 14:30
Raw Sample Received by: JH (WCL)
Raw Sample Relinquished by: BM CWL