



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: JIGNESH

Run Number: LB127103

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/31/2023	09:30
2	O4211-01		1	12	100.00	>212.0	>212.0	08/31/2023	10:00
3	O4211-01DUP		1	12	100.00	>212.0	>212.0	08/31/2023	10:30
4	O4215-01		1	14	100.00	>212.0	>212.0	08/31/2023	11:00
5	O4222-01		1	14	100.00	>212.0	>212.0	08/31/2023	11:30
6	O4222-02		1	12	100.00	>212.0	>212.0	08/31/2023	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

60127103

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-W8-23

WorkList ID : 173189

Department : Wet-Chemistry

Date : 08-30-2023 18:54:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4211-01	TRE-1767	Water	Flash Point	Cool 4 deg C	PSEG03	L12	08/29/2023	1010B
O4215-01	AVE-DRUM-04-20230828	Water	Flash Point	Cool 4 deg C	TETR16	L21	08/28/2023	1010B
O4222-01	1039	Water	Flash Point	Cool 4 deg C	PSEG03	N21	08/29/2023	1010B
O4222-02	1040	Water	Flash Point	Cool 4 deg C	PSEG03	N21	08/29/2023	1010B

Date/Time 08/31/2023 09:20
Raw Sample Received by: RM CWO
Raw Sample Relinquished by: MB (D)

Date/Time 08/31/2023 12:40
Raw Sample Received by: MB COW
Raw Sample Relinquished by: RM CWO