


LB127116

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.8095e-004 slope

C= -9.1962e-003 intercept

Rho= 0.9999954

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.00	0.000	0.019	0.019	156	0.000	156					
0.05	0.050					501	0				
0.20	0.200	0.191	-0.009	1105	0.0 %	1105					
2.50	2.500	2.497	-0.003	13849	0.0 %	13849					
5.00	5.000	4.987	-0.013	27610	0.0 %	27610					
7.50	7.500	7.497	-0.003	41480	0.0 %	41480					
10.0	10.000	10.010	0.010	55371	0.0 %	55371					

0.00511

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	156	-		0 SFAM01.1	PPB	9/1/2023 11:51	Std
	0.2 S01	1105	-		0.2 SFAM01.1	PPB	9/1/2023 11:54	Std
	2.5 S02	13849	-		2.5 SFAM01.1	PPB	9/1/2023 11:58	Std
	5 S03	27610	-		5 SFAM01.1	PPB	9/1/2023 12:00	Std
	7.5 S04	41480	-		7.5 SFAM01.1	PPB	9/1/2023 12:02	Std
	10 S05	55371	-		10 SFAM01.1	PPB	9/1/2023 12:04	Std
ICV044	ICV044	23184	4.1859 -		SFAM01.1	PPB	9/1/2023 12:07	SMPL
ICB044	ICB044	109	0.0105 -		SFAM01.1	PPB	9/1/2023 12:10	SMPL
CCV040	CCV040	27419	4.9522 -		SFAM01.1	PPB	9/1/2023 12:12	SMPL
CCB040	CCB040	101	0.0091 -		SFAM01.1	PPB	9/1/2023 12:14	SMPL
PB155298BL	PBS298	102	0.0093 -		SFAM01.1	PPB	9/1/2023 12:16	SMPL
O4040-01	MEZYJ6	2681	0.4759 -		SFAM01.1	PPB	9/1/2023 12:19	SMPL
O4040-02	MEZYJ7	2946	0.5239 -		SFAM01.1	PPB	9/1/2023 12:21	SMPL
O4040-03	MEZYJ7D	2591	0.4596 -		SFAM01.1	PPB	9/1/2023 12:26	SMPL
O4040-04	MEZYJ7S	16020	2.8896 -		SFAM01.1	PPB	9/1/2023 12:30	SMPL
O4040-05	MEZYJ8	1134	0.196 -		SFAM01.1	PPB	9/1/2023 12:32	SMPL
O4040-06	MEZYE1	2073	0.3659 -		SFAM01.1	PPB	9/1/2023 12:35	SMPL
O4040-07	MEZYE2	2123	0.375 -		SFAM01.1	PPB	9/1/2023 12:37	SMPL
O4040-08	MEZYE3	592	0.0979 -		SFAM01.1	PPB	9/1/2023 12:39	SMPL
O4040-09	MEZYW6	2017	0.3558 -		SFAM01.1	PPB	9/1/2023 12:42	SMPL
O4040-10	MEZYJ3	994	0.1707 -		SFAM01.1	PPB	9/1/2023 12:44	SMPL
O4040-11	MEZYJ4	485	0.0786 -		SFAM01.1	PPB	9/1/2023 12:46	SMPL
O4040-12	MEZYJ5	143	0.0167 -		SFAM01.1	PPB	9/1/2023 12:48	SMPL
O4040-13	MEZYH1	695	0.1166 -		SFAM01.1	PPB	9/1/2023 12:51	SMPL
O4040-14	MEZYH2	182	0.0237 -		SFAM01.1	PPB	9/1/2023 12:53	SMPL
O4040-15	MEZYH3	2206	0.39 -		SFAM01.1	PPB	9/1/2023 12:55	SMPL
O4040-16	MEZYD5	9381	1.6883 -		SFAM01.1	PPB	9/1/2023 12:57	SMPL
O4040-17	MEZYD6	19678	3.5515 -		SFAM01.1	PPB	9/1/2023 13:00	SMPL
O4040-18	MEZYD7	33189	5.9963 -		SFAM01.1	PPB	9/1/2023 13:02	SMPL
CCV041	CCV041	27376	4.9445 -		SFAM01.1	PPB	9/1/2023 13:04	SMPL
CCB041	CCB041	92	0.0075 -		SFAM01.1	PPB	9/1/2023 13:06	SMPL