

LB127122


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

61400



A= 0.0000e+000

B= 1.6511e-004

C= -1.9115e-002

Rho= 0.9995805

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.0	0.000	0.008	0.008	163	0.000	163					
0.2	0.200	0.165	-0.035	1113	0.0 %	1113					-18
2.5	2.500	2.632	0.132	16055	0.0 %	16055					5
5.0	5.000	4.962	-0.038	30171	0.0 %	30171					-1
7.5	7.500	7.315	-0.185	44417	0.0 %	44417					-2
10.0	10.000	10.119	0.119	61400	0.0 %	61400					1

LB127122 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Sep 2023 15:09:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	163	-	0.0000	7471B	PPB	01 Sep 2023 15:13:35	S	Std
0.2 - 1	50.2	1113	-	0.2000	7471B	PPB	01 Sep 2023 15:15:52	S	Std
2.5 - 1	52.5	16055	-	2.5000	7471B	PPB	01 Sep 2023 15:18:09	S	Std
5.0 - 1	55	30171	-	5.0000	7471B	PPB	01 Sep 2023 15:20:25	S	Std
7.5 - 1	57.5	44417	-	7.5000	7471B	PPB	01 Sep 2023 15:22:41	S	Std
10.0 - 1	51.0	61400	-	10.0000	7471B	PPB	01 Sep 2023 15:27:31	S	Std
ICV46 - 1	ICV46	25028	4.1133	-	7471B	PPB	01 Sep 2023 15:30:23	U	SMPL
ICB46 - 1	ICB46	98	-0.0029	-	7471B	PPB	01 Sep 2023 15:32:37	U	SMPL
CCV45 - 1	CCV45	29723	4.8885	-	7471B	PPB	01 Sep 2023 15:34:53	U	SMPL
CCB45 - 1	CCB45	104	-0.0019	-	7471B	PPB	01 Sep 2023 15:37:07	U	SMPL
CRA - 1	CRA	1214	0.1813	-	7471B	PPB	01 Sep 2023 15:45:47	U	SMPL
HighStd - 1	HighStd	61158	10.0788	-	7471B	PPB	01 Sep 2023 15:48:03	U	SMPL
ChkStd - 1	ChkStd	41999	6.9154	-	7471B	PPB	01 Sep 2023 15:50:20	U	SMPL
PB155295BL - 1	PBS	-24	-0.0231	-	7471B	PPB	01 Sep 2023 15:52:37	U	SMPL
PB155295BS - 1	LCSS	21928	3.6014	-	7471B	PPB	01 Sep 2023 15:54:53	U	SMPL
O4249-01 - 1	SU-02-083023	2177	0.3403	-	7471B	PPB	01 Sep 2023 15:57:09	U	SMPL
O4249-01DUP - 1	SU-02-083023DUP	1807	0.2792	-	7471B	PPB	01 Sep 2023 15:59:26	U	SMPL
O4249-01MS - 1	SU-02-083023MS	25774	4.2365	-	7471B	PPB	01 Sep 2023 16:01:42	U	SMPL
O4249-01MSD - 1	SU-02-083023MSD	25518	4.1942	-	7471B	PPB	01 Sep 2023 16:03:59	U	SMPL
O4257-01 - 1	CL-2-083123	2215	0.3466	-	7471B	PPB	01 Sep 2023 16:06:14	U	SMPL
CCV46 - 1	CCV46	29768	4.8959	-	7471B	PPB	01 Sep 2023 16:08:30	U	SMPL
CCB46 - 1	CCB46	78	-0.0062	-	7471B	PPB	01 Sep 2023 16:10:46	U	SMPL
O4264-01 - 1	VNJ-206	3291	0.5243	-	7471B	PPB	01 Sep 2023 16:13:02	U	SMPL
O4265-01 - 1	OR-3-083123	1999	0.3109	-	7471B	PPB	01 Sep 2023 16:15:17	U	SMPL
O4249-01LX5 - 1		395	0.0461	-	7471B	PPB	01 Sep 2023 16:17:32	U	SMPL
O4249-01A - 1		26685	4.3869	-	7471B	PPB	01 Sep 2023 16:19:48	U	SMPL
CCV47 - 1	CCV47	30319	4.9869	-	7471B	PPB	01 Sep 2023 16:22:04	U	SMPL
CCB47 - 1	CCB47	52	-0.0105	-	7471B	PPB	01 Sep 2023 16:24:20	U	SMPL