



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: JIGNESH

Run Number: LB127125

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/01/2023	14:50
2	O4232-11		1	12	100.00	>212.0	>212.0	09/01/2023	15:20
3	O4232-11DUP		1	13	100.00	>212.0	>212.0	09/01/2023	15:50
4	O4256-11		1	12	100.00	>212.0	>212.0	09/01/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127325

WorkList Name : FP-9-1

WorkList ID : 173240

Department : Wet-Chemistry

Date : 09-01-2023 10:59:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4232-11 D	50088	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/30/2023	1010B
O4256-11 S	Groundwater	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/31/2023	1010B

Date/Time 09/01/2023 14:40
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: MB (DL)

Date/Time 09/01/2023 16:30
 Raw Sample Received by: MB CDL
 Raw Sample Relinquished by: RM CWL