



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127133

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	09/05/2023	12:10
2	O4273-01		1	12	100.00	>212.0	>212.0	09/05/2023	12:40
3	O4273-01DUP		1	13	100.00	>212.0	>212.0	09/05/2023	13:10
4	O4274-03		1	14	100.00	>212.0	>212.0	09/05/2023	13:40
5	O4275-01		1	12	100.00	>212.0	>212.0	09/05/2023	14:10
6	O4278-01		1	12	100.00	>212.0	>212.0	09/05/2023	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16127133

WorkList Name : fp-9-05

WorkList ID : 173274

Department : Wet-Chemistry

Date : 09-05-2023 08:05:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4273-01F	1815	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/01/2023	1010B
O4274-03D	RB22282	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/01/2023	1010B
O4275-01C	60065	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/01/2023	1010B
O4278-01E	822-823	Water	Flash Point	Cool 4 deg C	PSEG03	N11	09/01/2023	1010B

Date/Time 09/05/2023 12:00
 Raw Sample Received by: RM LUG
 Raw Sample Relinquished by: JH (OC)

Date/Time 09/05/2023 15:30
 Raw Sample Received by: JH (OC)
 Raw Sample Relinquished by: RM LUG