

**PERCENT SOLID**

Supervisor: Iwona
Analyst: JIGNESH
Date: 9/7/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 09/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:44
Out Date: 09/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127143

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4289-01	TS-02-BLEND	20	1.15	8.81	9.96	8.92	88.2	
O4291-01	SOIL-CEMENT-COMP	1	1.15	8.37	9.52	7.72	78.5	
O4292-01	WASTE	2	1.16	8.65	9.81	8.6	86.0	
O4293-01	SOIL	3	1.12	8.58	9.7	8.22	82.8	
O4298-01	960	4	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4298-02	961	5	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4298-03	962	6	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4298-04	963	7	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O4299-01	SB-01-9.5-10.0	8	1.16	8.50	9.66	8.23	83.2	
O4299-02	SB-01-COMP	9	1.16	8.80	9.96	7.13	67.8	
O4299-03	SB-02-9.5-10.0	10	1.19	8.51	9.7	7.74	77.0	
O4299-04	SB-02-COMP	11	1.19	8.79	9.98	8.45	82.6	
O4299-05	SB-03-9.5-10.0	12	1.19	8.48	9.67	6.53	63.0	
O4299-06	SB-03-COMP	13	1.11	8.59	9.7	3.68	29.9	
O4299-07	SB-04-7.0-7.5	14	1.15	8.81	9.96	9.00	89.1	
O4299-08	SB-04-COMP	15	1.17	8.49	9.66	7.9	79.3	
O4304-04	WC-1	16	1.15	8.63	9.78	8.02	79.6	
O4304-05	WC-1	17	1.15	8.63	9.78	8.02	79.6	
O4305-01	TP-9-5-23-EPH	18	1.12	8.54	9.66	8.63	87.9	
O4305-02	TP-9-5-23-EPH	19	1.12	8.54	9.66	8.63	87.9	
O4312-01	RT4213	21	1.19	8.70	9.89	9.5	95.5	
O4312-02	RT4213-E2	22	1.12	8.40	9.52	9.33	97.7	
O4314-01	VNJ-201	23	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O4314-02	VNJ-201-E2	24	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O4321-01	RT2345	25	1.19	8.55	9.74	5.84	54.4	GEL MATRIX
O4321-02	RT2345-E2	26	1.16	8.49	9.65	8.6	87.6	GEL MATRIX
O4321-03	RT4296	27	1.17	8.56	9.73	8.76	88.7	
O4321-04	RT4296-E2	28	1.16	8.82	9.98	9.26	91.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$