



Analytical Summary Report

Analysis Method: SM3500-Cr B

ANALYST: Rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127144

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
Calibration Std. hexchrome 0.1 ppm	WP104009
Calibration Std. hexchrome 0.05 ppm	WP104008
calibration std. hexchrome 0.01 ppm	WP104006
calibration std. hexchrome 0 ppm	WP104005
hexavalent chromium color reagent	WP103967
0.2N SULFURIC ACID	WP103581
Calibration Std Hexachrome 0.025 ppm	WP104007
Hexavalent Chromium ICV-LCS Std	WP104012
Calibration and CCV std HexChrome 0.5PPM	WP104010
Calibration std HexChrome 1.0PPM	WP104011

Intercept: -0

Slope: 0.7816

Regression: 0.999993

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at 540nm		Absorbance Difference	Result (mg/L)	%D	Anal Date	Anal Time
								Backgrnd	Color					
1	CAL1	0	1	100	100		1.77	0.000	0.000	0.000	0		09/06/2023	10:00
2	CAL2	0.01	1	100	100		1.86	0.000	0.009	0.009	0.011	10	09/06/2023	10:01
3	CAL3	0.025	1	100	100		1.88	0.000	0.018	0.018	0.023	-8	09/06/2023	10:02
4	CAL4	0.05	1	100	100		1.86	0.000	0.038	0.038	0.048	-4	09/06/2023	10:03
5	CAL5	0.1	1	100	100		1.88	0.000	0.079	0.079	0.101	1	09/06/2023	10:04
6	CAL6	0.5	1	100	100		1.91	0.000	0.392	0.392	0.501	0.2	09/06/2023	10:05
7	CAL7	1	1	100	100		1.87	0.000	0.781	0.781	0.999	-0.1	09/06/2023	10:06



Analytical Summary Report

Analysis Method: SM3500-Cr B

ANALYST: Rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127144

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at540nm		Absorbance Difference	Intermediate Result (mg/L)	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		1.92	0.000	0.390	0.390	0.499	09/06/2023	10:07
2	ICB		1	100	100		1.79	0.000	0.000	0.000	0.000	09/06/2023	10:08
3	CCV1	0.5	1	100	100		1.93	0.000	0.392	0.392	0.502	09/06/2023	10:09
4	CCB1		1	100	100		1.80	0.000	0.000	0.000	0.000	09/06/2023	10:10
5	RL Check	0.01	1	100	100		1.91	0.000	0.009	0.009	0.012	09/06/2023	10:11
6	LB127144BL		1	100	100		1.80	0.000	0.000	0.000	0.000	09/06/2023	10:12
7	LB127144BS	0.5	1	100	100		1.90	0.0000	0.402	0.402	0.514	09/06/2023	10:13
8	O4299-09		1	100	100		2.04	0.004	0.004	0.000	0.000	09/06/2023	10:14
9	O4299-09DU		1	100	100		2.02	0.004	0.004	0.000	0.000	09/06/2023	10:15
10	O4299-09MS	1	2	100	100		2.06	0.003	0.362	0.359	0.459	09/06/2023	10:16
11	O4299-09MS	1	2	100	100		2.04	0.003	0.385	0.382	0.489	09/06/2023	10:17
12	CCV2	0.5	1	100	100		1.94	0.000	0.391	0.391	0.500	09/06/2023	10:18
13	CCB2		1	100	100		1.78	0.000	0.000	0.000	0.000	09/06/2023	10:19

WORKLIST(Hardcopy Internal Chain)

6127144

WorkList Name : hex 9-23

WorkList ID : 173307

Department : Wet-Chemistry

Date : 09-05-2023 12:40:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4299-09	MW-01	Water	Hexavalent Chromium	Ammonium sulfate buffer	LIRO01	N11	09/05/2023	SM3500-Cr B

Date/Time 09/05/2023 16:00
 Raw Sample Received by: RM
 Raw Sample Relinquished by: JH (CWC)

Date/Time 09/05/2023 16:50
 Raw Sample Received by: JH (CWC)
 Raw Sample Relinquished by: RM (CWC)