



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127146

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/06/2023	11:40
2	CAL2	1	Water	NA	NA	20.7	7.00	09/06/2023	11:41
3	CAL3	1	Water	NA	NA	20.7	10.02	09/06/2023	11:43
4	ICV	1	Water	NA	NA	20.6	6.98	09/06/2023	11:45
5	CCV1	1	Water	NA	NA	20.7	2.01	09/06/2023	11:48
6	O4293-01	1	Solid	20.02	20	22.9	8.24	09/06/2023	11:50
7	O4293-01DUP	1	Solid	20.03	20	22.8	8.25	09/06/2023	11:51
8	CCV2	1	Water	NA	NA	20.7	12.02	09/06/2023	12:00

WORKLIST(Hardcopy Internal Chain)

1731246

WorkList Name : ph 04293

WorkList ID : 173323

Department : Wet-Chemistry

Date : 09-06-2023 11:26:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4293-01	SOIL	Solid	pH	Cool 4 deg C	LETI02	N11	09/05/2023	9045D

Date/Time

09/06/23 11:35

Raw Sample Received by:

70100C1

Raw Sample Relinquished by:

70100C1

Date/Time

09/06/23

Raw Sample Received by:

70100C1

Raw Sample Relinquished by:

70100C1